

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX050222\
 Data File : VX028485.D
 Acq On : 02 May 2022 19:55
 Operator : JC/MD
 Sample : N2616-11
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MH-20D

Quant Time: May 03 04:54:13 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X041922W.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 19 13:38:04 2022
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	372487	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	654046	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	678131	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	314492	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	277330	57.101	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	114.200%
35) Dibromofluoromethane	5.391	113	246299	53.698	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	107.400%
50) Toluene-d8	8.653	98	945906	53.693	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	107.380%
62) 4-Bromofluorobenzene	11.085	95	348320	51.711	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	103.420%
Target Compounds						Qvalue
16) Acetone	2.373	43	10505	5.718	ug/l	88
20) Methylene Chloride	2.788	84	13029	3.616	ug/l	93
43) Isopropyl Acetate	6.348	43	26134	2.645	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX050222\
Data File : VX028485.D
Acq On : 02 May 2022 19:55
Operator : JC/MD
Sample : N2616-11
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MH-20D

Quant Time: May 03 04:54:13 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X041922W.M
Quant Title : SW846 8260
QLast Update : Tue Apr 19 13:38:04 2022
Response via : Initial Calibration

