

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX050223\
 Data File : VX035467.D
 Acq On : 02 May 2023 10:19
 Operator : JC/MD
 Sample : VX0502MBL01
 Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_X/MEOH
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0502MBL01

Quant Time: May 03 03:13:59 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X042523W.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 25 16:31:40 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.544	168	242243	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	393587	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	367992	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	180136	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	140426	48.050	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	96.100%
35) Dibromofluoromethane	5.379	113	123095	50.956	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	101.920%
50) Toluene-d8	8.647	98	464604	50.983	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	101.960%
62) 4-Bromofluorobenzene	11.079	95	182890	53.350	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	106.700%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX050223\
Data File : VX035467.D
Acq On : 02 May 2023 10:19
Operator : JC/MD
Sample : VX0502MBL01
Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_X/MEOH
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0502MBL01

Quant Time: May 03 03:13:59 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X042523W.M
Quant Title : SW846 8260
QLast Update : Tue Apr 25 16:31:40 2023
Response via : Initial Calibration

