

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX050223\
 Data File : VX035474.D
 Acq On : 02 May 2023 13:16
 Operator : JC/MD
 Sample : 02533-15
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 EB-01

Quant Time: May 03 03:15:33 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X042523W.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 25 16:31:40 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	228274	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	375115	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	345281	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	162763	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	132385	48.070	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	96.140%
35) Dibromofluoromethane	5.385	113	116508	50.604	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	101.200%
50) Toluene-d8	8.647	98	435327	50.123	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	100.240%
62) 4-Bromofluorobenzene	11.079	95	169009	51.729	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	103.460%
Target Compounds						
					Qvalue	
16) Acetone	2.392	43	8202	6.857	ug/l	94
20) Methylene Chloride	2.788	84	2221	0.810	ug/l	86
25) 2-Butanone	4.599	43	1554	0.890	ug/l #	82

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX050223\
Data File : VX035474.D
Acq On : 02 May 2023 13:16
Operator : JC/MD
Sample : 02533-15
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
EB-01

Quant Time: May 03 03:15:33 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X042523W.M
Quant Title : SW846 8260
QLast Update : Tue Apr 25 16:31:40 2023
Response via : Initial Calibration

