

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX050223\
 Data File : VX035478.D
 Acq On : 02 May 2023 14:49
 Operator : JC/MD
 Sample : 02533-20 2000X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 DUP-1

Quant Time: May 03 03:16:06 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X042523W.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 25 16:31:40 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	228800	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	377308	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	332910	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	163739	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	134872	48.861	ug/l	0.00
Spiked Amount 50.000	Range 78 - 117		Recovery =	97.720%		
35) Dibromofluoromethane	5.385	113	115032	49.673	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery =	99.340%		
50) Toluene-d8	8.646	98	438593	50.205	ug/l	0.00
Spiked Amount 50.000	Range 92 - 112		Recovery =	100.420%		
62) 4-Bromofluorobenzene	11.079	95	165740	50.433	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery =	100.860%		
Target Compounds						
27) cis-1,2-Dichloroethene	4.489	96	152875	53.887	ug/l	84
44) Trichloroethene	7.128	130	96186	33.660	ug/l	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX050223\
Data File : VX035478.D
Acq On : 02 May 2023 14:49
Operator : JC/MD
Sample : 02533-20 2000X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
DUP-1

Quant Time: May 03 03:16:06 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X042523W.M
Quant Title : SW846 8260
QLast Update : Tue Apr 25 16:31:40 2023
Response via : Initial Calibration

