

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX050223\
 Data File : VX035481.D
 Acq On : 02 May 2023 15:58
 Operator : JC/MD
 Sample : 02521-05
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 GMW-10

Quant Time: May 03 03:16:33 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X042523W.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 25 16:31:40 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	239956	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	398040	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	365850	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	172382	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	141956	49.036	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	98.080%
35) Dibromofluoromethane	5.385	113	119065	48.736	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	97.480%
50) Toluene-d8	8.647	98	474419	51.478	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	102.960%
62) 4-Bromofluorobenzene	11.079	95	179387	51.743	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	103.480%
Target Compounds						
					Qvalue	
16) Acetone	2.380	43	8025	6.383	ug/l	100
65) Chlorobenzene	10.080	112	2937	0.371	ug/l	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX050223\
Data File : VX035481.D
Acq On : 02 May 2023 15:58
Operator : JC/MD
Sample : 02521-05
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
GMW-10

Quant Time: May 03 03:16:33 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X042523W.M
Quant Title : SW846 8260
QLast Update : Tue Apr 25 16:31:40 2023
Response via : Initial Calibration

