

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX050223\
 Data File : VX035485.D
 Acq On : 02 May 2023 17:32
 Operator : JC/MD
 Sample : 02533-09
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 LEA-11S

Quant Time: May 03 03:17:18 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X042523W.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 25 16:31:40 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	228833	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	368449	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	329378	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	162803	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	133819	48.472	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	96.940%
35) Dibromofluoromethane	5.391	113	115022	50.862	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	101.720%
50) Toluene-d8	8.647	98	429054	50.294	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	100.580%
62) 4-Bromofluorobenzene	11.079	95	163967	51.093	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	102.180%
Target Compounds						
					Qvalue	
16) Acetone	2.386	43	3717	3.100	ug/l	91
27) cis-1,2-Dichloroethene	4.495	96	5446	1.919	ug/l	84
44) Trichloroethene	7.129	130	2942	1.054	ug/l	89

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX050223\
Data File : VX035485.D
Acq On : 02 May 2023 17:32
Operator : JC/MD
Sample : 02533-09
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
LEA-11S

Quant Time: May 03 03:17:18 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X042523W.M
Quant Title : SW846 8260
QLast Update : Tue Apr 25 16:31:40 2023
Response via : Initial Calibration

