

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX050223\
 Data File : VX035486.D
 Acq On : 02 May 2023 17:55
 Operator : JC/MD
 Sample : 02533-11
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 LEA-11I

Quant Time: May 03 03:17:29 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X042523W.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 25 16:31:40 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	215833	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	349426	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	318317	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	145525	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	128161	49.219	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	98.440%
35) Dibromofluoromethane	5.385	113	108179	50.441	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	100.880%
50) Toluene-d8	8.647	98	408381	50.477	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	100.960%
62) 4-Bromofluorobenzene	11.079	95	153212	50.341	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	100.680%
Target Compounds						
					Qvalue	
16) Acetone	2.392	43	4620	4.085	ug/l #	85
27) cis-1,2-Dichloroethene	4.489	96	8267	3.089	ug/l	83
44) Trichloroethene	7.123	130	168083	63.513	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX050223\
Data File : VX035486.D
Acq On : 02 May 2023 17:55
Operator : JC/MD
Sample : 02533-11
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
LEA-11I

Quant Time: May 03 03:17:29 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X042523W.M
Quant Title : SW846 8260
QLast Update : Tue Apr 25 16:31:40 2023
Response via : Initial Calibration

