

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX050223\
 Data File : VX035488.D
 Acq On : 02 May 2023 18:42
 Operator : JC/MD
 Sample : 02533-07
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 LEA-11D

Quant Time: May 03 03:17:51 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X042523W.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 25 16:31:40 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	226294	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	367948	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	337711	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	156866	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	133616	48.942	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	97.880%
35) Dibromofluoromethane	5.385	113	114273	50.600	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	101.200%
50) Toluene-d8	8.647	98	428518	50.300	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	100.600%
62) 4-Bromofluorobenzene	11.079	95	165095	51.515	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	103.020%
Target Compounds						
					Qvalue	
16) Acetone	2.392	43	5065	4.272	ug/l	98
44) Trichloroethene	7.129	130	1329	0.477	ug/l	70

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX050223\
Data File : VX035488.D
Acq On : 02 May 2023 18:42
Operator : JC/MD
Sample : 02533-07
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 24 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
LEA-11D

Quant Time: May 03 03:17:51 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X042523W.M
Quant Title : SW846 8260
QLast Update : Tue Apr 25 16:31:40 2023
Response via : Initial Calibration

