

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX050223\
 Data File : VX035493.D
 Acq On : 02 May 2023 20:37
 Operator : JC/MD
 Sample : IBLK
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 29 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 IBLK

Quant Time: May 03 03:19:11 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X042523W.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 25 16:31:40 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	213603	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	340304	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	316331	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	158115	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	123285	47.841	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	95.680%
35) Dibromofluoromethane	5.385	113	103842	49.716	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	99.440%
50) Toluene-d8	8.647	98	406255	51.560	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	103.120%
62) 4-Bromofluorobenzene	11.079	95	161567	54.509	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	109.020%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX050223\
Data File : VX035493.D
Acq On : 02 May 2023 20:37
Operator : JC/MD
Sample : IBLK
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 29 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
IBLK

Quant Time: May 03 03:19:11 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X042523W.M
Quant Title : SW846 8260
QLast Update : Tue Apr 25 16:31:40 2023
Response via : Initial Calibration

