

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX050224\
 Data File : VX041351.D
 Acq On : 02 May 2024 14:13
 Operator : JC/MD
 Sample : PB160645ZHE#01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB160645ZHE#01

Quant Time: May 03 01:38:17 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X043024W.M
 Quant Title : SW846 8260
 QLast Update : Wed May 01 02:38:17 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	161453	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.757	114	240896	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	226887	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	124367	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	89647	51.004	ug/l	0.00
Spiked Amount 50.000	Range 78 - 117		Recovery	=	102.000%	
35) Dibromofluoromethane	5.385	113	72567	44.452	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery	=	88.900%	
50) Toluene-d8	8.647	98	300170	58.094	ug/l	0.00
Spiked Amount 50.000	Range 92 - 112		Recovery	=	116.180%#	
62) 4-Bromofluorobenzene	11.079	95	111163	55.577	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery	=	111.160%	
Target Compounds						
						Qvalue
16) Acetone	2.386	43	39612	48.018	ug/l	94
18) Methyl Acetate	2.709	43	2736	1.521	ug/l	99
20) Methylene Chloride	2.782	84	5700	3.736	ug/l	97
43) Isopropyl Acetate	6.342	43	116497	31.811	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX050224\
Data File : VX041351.D
Acq On : 02 May 2024 14:13
Operator : JC/MD
Sample : PB160645ZHE#01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB160645ZHE#01

Quant Time: May 03 01:38:17 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X043024W.M
Quant Title : SW846 8260
QLast Update : Wed May 01 02:38:17 2024
Response via : Initial Calibration

