

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX050224\
 Data File : VX041352.D
 Acq On : 02 May 2024 14:37
 Operator : JC/MD
 Sample : PB160645ZHE#02
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB160645ZHE#02

Quant Time: May 03 01:38:45 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X043024W.M
 Quant Title : SW846 8260
 QLast Update : Wed May 01 02:38:17 2024
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	167540	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.757	114	246211	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	277404	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	137756	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	92343	50.629	ug/l	0.00
Spiked Amount 50.000	Range 78 - 117		Recovery	=	101.260%	
35) Dibromofluoromethane	5.385	113	77783	46.619	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery	=	93.240%	
50) Toluene-d8	8.647	98	315667	59.775	ug/l	0.00
Spiked Amount 50.000	Range 92 - 112		Recovery	=	119.540%#	
62) 4-Bromofluorobenzene	11.079	95	126930	62.090	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery	=	124.180%#	
Target Compounds						
						Qvalue
16) Acetone	2.392	43	41258	48.211	ug/l	95
18) Methyl Acetate	2.715	43	3031	1.624	ug/l	98
20) Methylene Chloride	2.788	84	6078	3.839	ug/l #	90
43) Isopropyl Acetate	6.342	43	106676	28.501	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX050224\
Data File : VX041352.D
Acq On : 02 May 2024 14:37
Operator : JC/MD
Sample : PB160645ZHE#02
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB160645ZHE#02

Quant Time: May 03 01:38:45 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X043024W.M
Quant Title : SW846 8260
QLast Update : Wed May 01 02:38:17 2024
Response via : Initial Calibration

