

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX050224\
 Data File : VX041354.D
 Acq On : 02 May 2024 15:23
 Operator : JC/MD
 Sample : P2352-01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TB-01-050124

Quant Time: May 03 01:39:38 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X043024W.M
 Quant Title : SW846 8260
 QLast Update : Wed May 01 02:38:17 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	181509	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.763	114	231321	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	203012	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	100123	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	89703	45.397	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	90.800%
35) Dibromofluoromethane	5.385	113	79907	50.974	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	101.940%
50) Toluene-d8	8.647	98	281246	56.685	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	113.360%#
62) 4-Bromofluorobenzene	11.079	95	86964	45.278	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	90.560%
Target Compounds						
18) Methyl Acetate	2.709	43	4698	2.324	ug/l	Qvalue 100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX050224\
Data File : VX041354.D
Acq On : 02 May 2024 15:23
Operator : JC/MD
Sample : P2352-01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TB-01-050124

Quant Time: May 03 01:39:38 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X043024W.M
Quant Title : SW846 8260
QLast Update : Wed May 01 02:38:17 2024
Response via : Initial Calibration

