



Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX050224\
Data File : VX041339.D
Acq On : 02 May 2024 08:19
Operator : JC/MD
Sample : BFB
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X043024W.M
Title : SW846 8260
Last Update : Wed May 01 02:38:17 2024



AutoFind: Scans 1638, 1639, 1640; Background Corrected with Scan 1632

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	19.4	9171	PASS
75	95	30	60	51.9	24539	PASS
95	95	100	100	100.0	47323	PASS
96	95	5	9	6.9	3253	PASS
173	174	0.00	2	1.4	649	PASS
174	95	50	100	97.2	46008	PASS
175	174	5	9	7.6	3482	PASS
176	174	95	101	96.4	44339	PASS
177	176	5	9	6.7	2974	PASS