

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX050225\
 Data File : VX046021.D
 Acq On : 02 May 2025 12:41
 Operator : JC/MD
 Sample : Q1922-04
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MH-R

Quant Time: May 02 23:13:51 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X040225W.M
 Quant Title : SW846 8260
 QLast Update : Wed Apr 02 03:11:43 2025
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	61759	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	126035	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.049	117	119212	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	52343	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.946	65	62632	55.455	ug/l	0.00
Spiked Amount	50.000	Range 74 - 125	Recovery	=	110.920%	
35) Dibromofluoromethane	5.379	113	44822	50.124	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	=	100.240%	
50) Toluene-d8	8.647	98	158215	50.691	ug/l	0.00
Spiked Amount	50.000	Range 86 - 113	Recovery	=	101.380%	
62) 4-Bromofluorobenzene	11.079	95	61749	54.314	ug/l	0.00
Spiked Amount	50.000	Range 77 - 121	Recovery	=	108.620%	
Target Compounds						
					Qvalue	
16) Acetone	2.386	43	22192	47.852	ug/l	98
18) Methyl Acetate	2.703	43	1475	1.381	ug/l	95
20) Methylene Chloride	2.782	84	6362	7.328	ug/l	95
43) Isopropyl Acetate	6.348	43	11676	5.063	ug/l	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX050225\
Data File : VX046021.D
Acq On : 02 May 2025 12:41
Operator : JC/MD
Sample : Q1922-04
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MH-R

Quant Time: May 02 23:13:51 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X040225W.M
Quant Title : SW846 8260
QLast Update : Wed Apr 02 03:11:43 2025
Response via : Initial Calibration

