

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX050225\
 Data File : VX046033.D
 Acq On : 02 May 2025 17:22
 Operator : JC/MD
 Sample : Q1932-08
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 COMP-7

Quant Time: May 02 23:19:54 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X040225W.M
 Quant Title : SW846 8260
 QLast Update : Wed Apr 02 03:11:43 2025
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	65765	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	128974	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.049	117	124796	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	53600	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	64991	54.039	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 108.080%	
35) Dibromofluoromethane	5.385	113	46078	50.355	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 100.700%	
50) Toluene-d8	8.647	98	166270	52.058	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 104.120%	
62) 4-Bromofluorobenzene	11.079	95	65807	56.564	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	= 113.120%	
Target Compounds						
					Qvalue	
16) Acetone	2.386	43	20744	42.005	ug/l	97
18) Methyl Acetate	2.709	43	1965	1.728	ug/l	98
20) Methylene Chloride	2.782	84	2776	3.003	ug/l	97
43) Isopropyl Acetate	6.342	43	11728	4.970	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX050225\
Data File : VX046033.D
Acq On : 02 May 2025 17:22
Operator : JC/MD
Sample : Q1932-08
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
COMP-7

Quant Time: May 02 23:19:54 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X040225W.M
Quant Title : SW846 8260
QLast Update : Wed Apr 02 03:11:43 2025
Response via : Initial Calibration

