

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX050319\
 Data File : VX009317.D
 Acq On : 03 May 2019 18:14
 Operator : JC/SP
 Sample : K2662-03
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 SS050MW608-190430

Quant Time: May 04 00:06:24 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X042919W.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 30 07:03:50 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.66	168	195539	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	315189	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	275764	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	119630	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.06	65	128443	48.12	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.24%	
35) Dibromofluoromethane	5.50	113	94802	47.85	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.70%	
50) Toluene-d8	8.71	98	392500	49.07	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.14%	
62) 4-Bromofluorobenzene	11.13	95	131324	45.18	ug/l	0.00
Spiked Amount 50.000			Recovery	=	90.36%	
Target Compounds						
27) cis-1,2-Dichloroethene	4.59	96	8482	3.571	ug/l	91
44) Trichloroethene	7.21	130	13644	6.226	ug/l	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX050319\
Data File : VX009317.D
Acq On : 03 May 2019 18:14
Operator : JC/SP
Sample : K2662-03
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
SS050MW608-190430

Quant Time: May 04 00:06:24 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X042919W.M
Quant Title : SW846 8260
QLast Update : Tue Apr 30 07:03:50 2019
Response via : Initial Calibration

