

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX050319\
 Data File : VX009325.D
 Acq On : 03 May 2019 21:20
 Operator : JC/SP
 Sample : K2687-05
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 27 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 DUP-0520-190502

Quant Time: May 04 06:40:13 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X042919W.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 30 07:03:50 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	197442	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	319111	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	280230	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	118769	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.06	65	132158	49.03	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.06%	
35) Dibromofluoromethane	5.50	113	97226	48.47	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.94%	
50) Toluene-d8	8.71	98	399581	49.34	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.68%	
62) 4-Bromofluorobenzene	11.13	95	133993	45.53	ug/l	0.00
Spiked Amount 50.000			Recovery	=	91.06%	
Target Compounds						
19) Methyl tert-butyl Ether	3.20	73	30426	4.229	ug/l	95
27) cis-1,2-Dichloroethene	4.59	96	16272	6.785	ug/l	91
44) Trichloroethene	7.21	130	7770	3.502	ug/l	96
64) Tetrachloroethene	9.34	164	4324	2.249	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX050319\
Data File : VX009325.D
Acq On : 03 May 2019 21:20
Operator : JC/SP
Sample : K2687-05
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 27 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
DUP-0520-190502

Quant Time: May 04 06:40:13 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X042919W.M
Quant Title : SW846 8260
QLast Update : Tue Apr 30 07:03:50 2019
Response via : Initial Calibration

