

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX050319\
 Data File : VX009328.D
 Acq On : 03 May 2019 22:30
 Operator : JC/SP
 Sample : K2687-08
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 30 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 SS052OW377-190502

Quant Time: May 04 06:48:35 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X042919W.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 30 07:03:50 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	197576	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	320531	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	280489	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	120537	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.06	65	132794	49.23	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.46%	
35) Dibromofluoromethane	5.50	113	98456	48.86	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.72%	
50) Toluene-d8	8.71	98	398810	49.02	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.04%	
62) 4-Bromofluorobenzene	11.13	95	133587	45.19	ug/l	0.00
Spiked Amount 50.000			Recovery	=	90.38%	
Target Compounds						
					Qvalue	
3) Chloromethane	1.32	50	1226	0.493	ug/l #	80
16) Acetone	2.45	43	4390	3.031	ug/l	94
27) cis-1,2-Dichloroethene	4.60	96	3906	1.628	ug/l	87
30) Chloroform	5.22	83	2065	0.533	ug/l	95
44) Trichloroethene	7.21	130	10797	4.845	ug/l	89
64) Tetrachloroethene	9.34	164	3226	1.677	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX050319\
Data File : VX009328.D
Acq On : 03 May 2019 22:30
Operator : JC/SP
Sample : K2687-08
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 30 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
SS052OW377-190502

Quant Time: May 04 06:48:35 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X042919W.M
Quant Title : SW846 8260
QLast Update : Tue Apr 30 07:03:50 2019
Response via : Initial Calibration

