

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX050319\
 Data File : VX009329.D
 Acq On : 03 May 2019 22:53
 Operator : JC/SP
 Sample : K2687-09
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 31 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 SS052MW718-190502

Quant Time: May 06 04:59:14 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X042919W.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 30 07:03:50 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	198140	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	319469	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	279985	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	121299	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.06	65	131583	48.65	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.30%	
35) Dibromofluoromethane	5.50	113	97190	48.39	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.78%	
50) Toluene-d8	8.71	98	399806	49.31	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.62%	
62) 4-Bromofluorobenzene	11.13	95	132961	45.13	ug/l	0.00
Spiked Amount 50.000			Recovery	=	90.26%	
Target Compounds						
16) Acetone	2.44	43	4721	3.250	ug/l	92
27) cis-1,2-Dichloroethene	4.60	96	11367	4.723	ug/l	86
44) Trichloroethene	7.21	130	27825	12.527	ug/l	98
55) 1,1,2-Trichloroethane	9.22	97	803	0.352	ug/l	92
64) Tetrachloroethene	9.34	164	3888	2.024	ug/l #	79

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX050319\
Data File : VX009329.D
Acq On : 03 May 2019 22:53
Operator : JC/SP
Sample : K2687-09
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 31 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
SS052MW718-190502

Quant Time: May 06 04:59:14 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X042919W.M
Quant Title : SW846 8260
QLast Update : Tue Apr 30 07:03:50 2019
Response via : Initial Calibration

