

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX050319\
 Data File : VX009330.D
 Acq On : 03 May 2019 23:17
 Operator : JC/SP
 Sample : K2687-10
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 32 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 Z4-TB02-190502

Quant Time: May 04 01:30:19 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X042919W.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 30 07:03:50 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.66	168	190274	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	312482	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	270016	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	115862	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.06	65	128337	49.41	ug/l	0.00
Spiked Amount	50.000		Recovery	=	98.82%	
35) Dibromofluoromethane	5.50	113	95131	48.43	ug/l	0.00
Spiked Amount	50.000		Recovery	=	96.86%	
50) Toluene-d8	8.71	98	388292	48.96	ug/l	0.00
Spiked Amount	50.000		Recovery	=	97.92%	
62) 4-Bromofluorobenzene	11.13	95	127604	44.28	ug/l	0.00
Spiked Amount	50.000		Recovery	=	88.56%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX050319\
Data File : VX009330.D
Acq On : 03 May 2019 23:17
Operator : JC/SP
Sample : K2687-10
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 32 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
Z4-TB02-190502

Quant Time: May 04 01:30:19 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X042919W.M
Quant Title : SW846 8260
QLast Update : Tue Apr 30 07:03:50 2019
Response via : Initial Calibration

