

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX050322\
 Data File : VX028493.D
 Acq On : 03 May 2022 11:45
 Operator : JC/MD
 Sample : VX0503WBL01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0503WBL01

Quant Time: May 04 06:02:46 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X041922W.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 19 13:38:04 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	385184	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	674797	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	654756	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	301216	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	273612	54.479	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	108.960%
35) Dibromofluoromethane	5.385	113	249446	52.711	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	105.420%
50) Toluene-d8	8.653	98	923012	50.782	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	101.560%
62) 4-Bromofluorobenzene	11.079	95	329197	47.369	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	94.740%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX050322\
Data File : VX028493.D
Acq On : 03 May 2022 11:45
Operator : JC/MD
Sample : VX0503WBL01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0503WBL01

Quant Time: May 04 06:02:46 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X041922W.M
Quant Title : SW846 8260
QLast Update : Tue Apr 19 13:38:04 2022
Response via : Initial Calibration

