

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX050322\
 Data File : VX028498.D
 Acq On : 03 May 2022 13:56
 Operator : JC/MD
 Sample : N2642-01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 30988

Quant Time: May 04 06:04:31 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X041922W.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 19 13:38:04 2022
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	393632	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	676466	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	671514	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	341569	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	272487	53.090	ug/l	0.00
Spiked Amount 50.000	Range	61 - 141	Recovery	= 106.180%		
35) Dibromofluoromethane	5.391	113	251890	53.096	ug/l	0.00
Spiked Amount 50.000	Range	69 - 133	Recovery	= 106.200%		
50) Toluene-d8	8.653	98	956222	52.480	ug/l	0.00
Spiked Amount 50.000	Range	65 - 126	Recovery	= 104.960%		
62) 4-Bromofluorobenzene	11.079	95	361784	51.929	ug/l	0.00
Spiked Amount 50.000	Range	58 - 135	Recovery	= 103.860%		
Target Compounds						
						Qvalue
11) Tert butyl alcohol	2.959	59	29028	20.409	ug/l	96
16) Acetone	2.374	43	1027424	657.676	ug/l	96
19) Methyl tert-butyl Ether	3.111	73	3852	0.305	ug/l	99
20) Methylene Chloride	2.794	84	3260	0.856	ug/l #	79
25) 2-Butanone	4.556	43	352689	116.029	ug/l	99
29) Tetrahydrofuran	5.001	42	218918	110.455	ug/l	99
30) Chloroform	5.105	83	7764	0.947	ug/l	90
40) Benzene	6.038	78	11434	0.631	ug/l	99
51) 4-Methyl-2-Pentanone	8.574	43	99264	15.137	ug/l	97
52) Toluene	8.720	92	297177	23.615	ug/l	99
59) 2-Hexanone	9.433	43	9841	1.937	ug/l	90
65) Chlorobenzene	10.080	112	5300	0.370	ug/l	89
67) Ethyl Benzene	10.195	91	11147	0.464	ug/l	100
68) m/p-Xylenes	10.305	106	19545	2.000	ug/l	93
69) o-Xylene	10.640	106	20941	2.150	ug/l	97
80) 1,3,5-Trimethylbenzene	11.451	105	8480	0.426	ug/l	97
84) 1,2,4-Trimethylbenzene	11.756	105	34172	1.725	ug/l	93
95) Naphthalene	13.774	128	110347	4.880	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX050322\
Data File : VX028498.D
Acq On : 03 May 2022 13:56
Operator : JC/MD
Sample : N2642-01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
30988

Quant Time: May 04 06:04:31 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X041922W.M
Quant Title : SW846 8260
QLast Update : Tue Apr 19 13:38:04 2022
Response via : Initial Calibration

