

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX050322\
 Data File : VX028502.D
 Acq On : 03 May 2022 15:29
 Operator : JC/MD
 Sample : N2665-14
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TB

Quant Time: May 04 06:05:41 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X041922W.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 19 13:38:04 2022
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	362577	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	636948	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	631995	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	294539	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	262773	55.583	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	111.160%
35) Dibromofluoromethane	5.391	113	234968	52.602	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	105.200%
50) Toluene-d8	8.653	98	886802	51.689	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	103.380%
62) 4-Bromofluorobenzene	11.085	95	318775	48.595	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	97.180%
Target Compounds						
20) Methylene Chloride	2.782	84	1245	0.355	ug/l	Qvalue # 82

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX050322\
Data File : VX028502.D
Acq On : 03 May 2022 15:29
Operator : JC/MD
Sample : N2665-14
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TB

Quant Time: May 04 06:05:41 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X041922W.M
Quant Title : SW846 8260
QLast Update : Tue Apr 19 13:38:04 2022
Response via : Initial Calibration

