

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX050322\
 Data File : VX028503.D
 Acq On : 03 May 2022 15:52
 Operator : JC/MD
 Sample : PB144525TB
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB144525TB

Quant Time: May 04 06:05:54 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X041922W.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 19 13:38:04 2022
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	362475	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	639474	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	666061	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	310452	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	270314	57.194	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	114.380%
35) Dibromofluoromethane	5.391	113	243055	54.198	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	108.400%
50) Toluene-d8	8.653	98	934888	54.277	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	108.560%
62) 4-Bromofluorobenzene	11.085	95	345356	52.439	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	104.880%
Target Compounds						Qvalue
16) Acetone	2.379	43	9936	5.518	ug/l	94
20) Methylene Chloride	2.794	84	8864	2.528	ug/l	94
43) Isopropyl Acetate	6.348	43	26231	2.715	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX050322\
Data File : VX028503.D
Acq On : 03 May 2022 15:52
Operator : JC/MD
Sample : PB144525TB
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB144525TB

Quant Time: May 04 06:05:54 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X041922W.M
Quant Title : SW846 8260
QLast Update : Tue Apr 19 13:38:04 2022
Response via : Initial Calibration

