

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX050322\
 Data File : VX028504.D
 Acq On : 03 May 2022 16:16
 Operator : JC/MD
 Sample : PB144525ZHE#01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB144525ZHE#01

Quant Time: May 04 06:06:09 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X041922W.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 19 13:38:04 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	356289	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	629463	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	650232	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	302476	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	263594	56.741	ug/l	0.00
Spiked Amount 50.000	Range 61 - 141		Recovery = 113.480%			
35) Dibromofluoromethane	5.391	113	237665	53.839	ug/l	0.00
Spiked Amount 50.000	Range 69 - 133		Recovery = 107.680%			
50) Toluene-d8	8.653	98	912349	53.811	ug/l	0.00
Spiked Amount 50.000	Range 65 - 126		Recovery = 107.620%			
62) 4-Bromofluorobenzene	11.085	95	331851	51.190	ug/l	0.00
Spiked Amount 50.000	Range 58 - 135		Recovery = 102.380%			
Target Compounds					Qvalue	
16) Acetone	2.379	43	9752	5.508	ug/l	99
20) Methylene Chloride	2.788	84	8695	2.523	ug/l	95
43) Isopropyl Acetate	6.342	43	25519	2.684	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX050322\
Data File : VX028504.D
Acq On : 03 May 2022 16:16
Operator : JC/MD
Sample : PB144525ZHE#01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB144525ZHE#01

Quant Time: May 04 06:06:09 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X041922W.M
Quant Title : SW846 8260
QLast Update : Tue Apr 19 13:38:04 2022
Response via : Initial Calibration

