

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX050322\
 Data File : VX028505.D
 Acq On : 03 May 2022 16:39
 Operator : JC/MD
 Sample : PB144525ZHE#02
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB144525ZHE#02

Quant Time: May 04 06:06:28 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X041922W.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 19 13:38:04 2022
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	359422	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	637995	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	666053	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	313932	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	272116	58.064	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	116.120%
35) Dibromofluoromethane	5.391	113	243596	54.444	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	108.880%
50) Toluene-d8	8.653	98	938720	54.626	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	109.260%
62) 4-Bromofluorobenzene	11.085	95	346314	52.706	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	105.420%
Target Compounds						
						Qvalue
16) Acetone	2.380	43	10033	5.645	ug/l	98
20) Methylene Chloride	2.794	84	9638	2.772	ug/l #	83
43) Isopropyl Acetate	6.348	43	26339	2.733	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX050322\
Data File : VX028505.D
Acq On : 03 May 2022 16:39
Operator : JC/MD
Sample : PB144525ZHE#02
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB144525ZHE#02

Quant Time: May 04 06:06:28 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X041922W.M
Quant Title : SW846 8260
QLast Update : Tue Apr 19 13:38:04 2022
Response via : Initial Calibration

