

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX050322\
 Data File : VX028506.D
 Acq On : 03 May 2022 17:02
 Operator : JC/MD
 Sample : PB144525ZHE#03
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB144525ZHE#03

Quant Time: May 04 06:06:44 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X041922W.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 19 13:38:04 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	369964	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	652427	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	672438	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	310626	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	276328	57.283	ug/l	0.00
Spiked Amount 50.000	Range 61 - 141		Recovery = 114.560%			
35) Dibromofluoromethane	5.391	113	247933	54.188	ug/l	0.00
Spiked Amount 50.000	Range 69 - 133		Recovery = 108.380%			
50) Toluene-d8	8.653	98	952842	54.221	ug/l	0.00
Spiked Amount 50.000	Range 65 - 126		Recovery = 108.440%			
62) 4-Bromofluorobenzene	11.079	95	348872	51.921	ug/l	0.00
Spiked Amount 50.000	Range 58 - 135		Recovery = 103.840%			
Target Compounds					Qvalue	
16) Acetone	2.380	43	10597	5.829	ug/l	93
20) Methylene Chloride	2.794	84	9861	2.755	ug/l	91
43) Isopropyl Acetate	6.348	43	26454	2.684	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX050322\
Data File : VX028506.D
Acq On : 03 May 2022 17:02
Operator : JC/MD
Sample : PB144525ZHE#03
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB144525ZHE#03

Quant Time: May 04 06:06:44 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X041922W.M
Quant Title : SW846 8260
QLast Update : Tue Apr 19 13:38:04 2022
Response via : Initial Calibration

