

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX050322\
 Data File : VX028507.D
 Acq On : 03 May 2022 17:26
 Operator : JC/MD
 Sample : PB144525ZHE#04
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB144525ZHE#04

Quant Time: May 04 06:07:03 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X041922W.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 19 13:38:04 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	364189	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	646529	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	674326	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	314133	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	274347	57.774	ug/l	0.00
Spiked Amount 50.000	Range 61 - 141		Recovery = 115.540%			
35) Dibromofluoromethane	5.391	113	249976	55.133	ug/l	0.00
Spiked Amount 50.000	Range 69 - 133		Recovery = 110.260%			
50) Toluene-d8	8.653	98	952049	54.670	ug/l	0.00
Spiked Amount 50.000	Range 65 - 126		Recovery = 109.340%			
62) 4-Bromofluorobenzene	11.085	95	351511	52.791	ug/l	0.00
Spiked Amount 50.000	Range 58 - 135		Recovery = 105.580%			
Target Compounds						
16) Acetone	2.380	43	10413	5.816	ug/l	91
20) Methylene Chloride	2.788	84	9392	2.666	ug/l	97
43) Isopropyl Acetate	6.348	43	26890	2.753	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX050322\
Data File : VX028507.D
Acq On : 03 May 2022 17:26
Operator : JC/MD
Sample : PB144525ZHE#04
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB144525ZHE#04

Quant Time: May 04 06:07:03 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X041922W.M
Quant Title : SW846 8260
QLast Update : Tue Apr 19 13:38:04 2022
Response via : Initial Calibration

