

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX050322\
 Data File : VX028508.D
 Acq On : 03 May 2022 17:49
 Operator : JC/MD
 Sample : N2641-03
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TP-21

Quant Time: May 04 06:07:17 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X041922W.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 19 13:38:04 2022
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	364176	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	646676	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	661255	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	306281	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	273119	57.518	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	115.040%
35) Dibromofluoromethane	5.391	113	242627	53.500	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	107.000%
50) Toluene-d8	8.653	98	934386	53.644	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	107.280%
62) 4-Bromofluorobenzene	11.079	95	343536	51.582	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	103.160%
Target Compounds						
						Qvalue
16) Acetone	2.374	43	13049	7.644	ug/l	89
20) Methylene Chloride	2.788	84	10922	3.100	ug/l	98
43) Isopropyl Acetate	6.355	43	25554	2.616	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX050322\
Data File : VX028508.D
Acq On : 03 May 2022 17:49
Operator : JC/MD
Sample : N2641-03
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TP-21

Quant Time: May 04 06:07:17 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X041922W.M
Quant Title : SW846 8260
QLast Update : Tue Apr 19 13:38:04 2022
Response via : Initial Calibration

