

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX050322\
 Data File : VX028511.D
 Acq On : 03 May 2022 18:59
 Operator : JC/MD
 Sample : N2620-01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 SOIL-IDW

Quant Time: May 04 06:08:01 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X041922W.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 19 13:38:04 2022
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	360848	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	635239	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	657618	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	308038	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	263966	56.103	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	112.200%
35) Dibromofluoromethane	5.391	113	236334	53.051	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	106.100%
50) Toluene-d8	8.653	98	902529	52.748	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	105.500%
62) 4-Bromofluorobenzene	11.085	95	329047	50.296	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	100.600%
Target Compounds						
						Qvalue
16) Acetone	2.373	43	13202	7.835	ug/l	88
20) Methylene Chloride	2.788	84	11443	3.278	ug/l	97
43) Isopropyl Acetate	6.348	43	22018	2.294	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX050322\
Data File : VX028511.D
Acq On : 03 May 2022 18:59
Operator : JC/MD
Sample : N2620-01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
SOIL-IDW

Quant Time: May 04 06:08:01 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X041922W.M
Quant Title : SW846 8260
QLast Update : Tue Apr 19 13:38:04 2022
Response via : Initial Calibration

