

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX050322\
 Data File : VX028513.D
 Acq On : 03 May 2022 19:45
 Operator : JC/MD
 Sample : N2602-03
 Misc : 5.02g/10mL/100uL/5.00mL/MSVOA_X/MEOH
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 12-15-SB4-BK-PRISON

Quant Time: May 04 06:08:49 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X041922W.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 19 13:38:04 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	423142	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	701606	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	625827	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	287923	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	237781	43.097	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	86.200%
35) Dibromofluoromethane	5.391	113	213536	43.399	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	86.800%
50) Toluene-d8	8.653	98	804525	42.572	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	85.140%
62) 4-Bromofluorobenzene	11.085	95	273274	37.819	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	75.640%
Target Compounds						
					Qvalue	
18) Methyl Acetate	2.739	43	5297	1.080	ug/l	99
20) Methylene Chloride	2.782	84	1835	0.448	ug/l #	89
25) 2-Butanone	4.599	43	26127	7.996	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX050322\
Data File : VX028513.D
Acq On : 03 May 2022 19:45
Operator : JC/MD
Sample : N2602-03
Misc : 5.02g/10mL/100uL/5.00mL/MSVOA_X/MEOH
ALS Vial : 24 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
12-15-SB4-BK-PRISON

Quant Time: May 04 06:08:49 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X041922W.M
Quant Title : SW846 8260
QLast Update : Tue Apr 19 13:38:04 2022
Response via : Initial Calibration

