

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX050322\
 Data File : VX028515.D
 Acq On : 03 May 2022 20:32
 Operator : JC/MD
 Sample : N2643-03
 Misc : 1.20g/10mL/100uL/5.00mL/MSVOA_X/MEOH
 ALS Vial : 26 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 31487

Quant Time: May 04 06:09:23 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X041922W.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 19 13:38:04 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	491171	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	793238	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	753005	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.030	152	425620	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	258283	40.330	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	80.660%
35) Dibromofluoromethane	5.391	113	242386	43.572	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	87.140%
50) Toluene-d8	8.653	98	922565	43.179	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	86.360%
62) 4-Bromofluorobenzene	11.085	95	397195	48.619	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	97.240%
Target Compounds						
					Qvalue	
16) Acetone	2.428	43	20196	8.979	ug/l	94
25) 2-Butanone	4.611	43	29422	7.757	ug/l	98
31) Cyclohexane	5.489	56	623313	79.885	ug/l #	43
39) Methylcyclohexane	7.379	83	486707	53.159	ug/l	95
52) Toluene	8.732	92	18036841	1222.269	ug/l #	62
67) Ethyl Benzene	10.195	91	791739	29.410	ug/l	96
68) m/p-Xylenes	10.305	106	1122366	102.438	ug/l	96
69) o-Xylene	10.647	106	594035	54.379	ug/l	96
73) Isopropylbenzene	10.964	105	375954	12.851	ug/l	99
78) n-propylbenzene	11.311	91	1494233	46.063	ug/l	98
80) 1,3,5-Trimethylbenzene	11.457	105	1775850	71.513	ug/l	96
84) 1,2,4-Trimethylbenzene	11.756	105	6306729	255.489	ug/l	91
85) sec-Butylbenzene	11.896	105	921662	30.127	ug/l	82
86) p-Isopropyltoluene	12.012	119	559144	21.186	ug/l	98
89) n-Butylbenzene	12.335	91	1984192	91.559	ug/l #	51
95) Naphthalene	13.780	128	848642	30.118	ug/l #	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX050322\
Data File : VX028515.D
Acq On : 03 May 2022 20:32
Operator : JC/MD
Sample : N2643-03
Misc : 1.20g/10mL/100uL/5.00mL/MSVOA_X/MEOH
ALS Vial : 26 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
31487

Quant Time: May 04 06:09:23 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X041922W.M
Quant Title : SW846 8260
QLast Update : Tue Apr 19 13:38:04 2022
Response via : Initial Calibration

