

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX050322\
 Data File : VX028491.D
 Acq On : 03 May 2022 10:47
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: May 04 06:02:08 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X041922W.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 19 13:38:04 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	104	0.00
2 T	Dichlorodifluoromethane	0.323	0.417	-29.1#	151#	0.00
3 P	Chloromethane	0.375	0.401	-6.9	128	0.00
4 C	Vinyl Chloride	0.504	0.489	3.0#	113	0.00
5 T	Bromomethane	0.522	0.497	4.8	106	0.00
6 T	Chloroethane	0.461	0.356	22.8	104	0.00
7 T	Trichlorofluoromethane	1.112	0.927	16.6	96	0.00
8 T	Diethyl Ether	0.382	0.361	5.5	112	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.405	0.499	-23.2	159#	0.00
10 T	Methyl Iodide	0.509	0.653	-28.3#	137	0.00
11 T	Tert butyl alcohol	0.135	0.123	8.9	110	0.00
12 CM	1,1-Dichloroethene	0.384	0.446	-16.1#	148	0.00
13 T	Acrolein	0.080	0.077	3.8	92	0.00
14 T	Allyl chloride	0.645	0.656	-1.7	115	0.00
15 T	Acrylonitrile	0.266	0.295	-10.9	119	0.00
16 T	Acetone	0.215	0.238	-10.7	135	0.00
17 T	Carbon Disulfide	1.035	0.903	12.8	102	0.00
18 T	Methyl Acetate	0.580	0.641	-10.5	128	0.00
19 T	Methyl tert-butyl Ether	1.602	1.763	-10.0	115	0.00
20 T	Methylene Chloride	0.484	0.546	-12.8	134	0.00
21 T	trans-1,2-Dichloroethene	0.523	0.520	0.6	108	0.00
22 T	Diisopropyl ether	1.469	1.534	-4.4	113	0.00
23 T	Vinyl Acetate	1.283	1.335	-4.1	109	0.00
24 P	1,1-Dichloroethane	0.867	0.924	-6.6	122	0.00
25 T	2-Butanone	0.386	0.392	-1.6	110	0.00
26 T	2,2-Dichloropropane	0.847	0.786	7.2	101	0.00
27 T	cis-1,2-Dichloroethene	0.623	0.652	-4.7	115	0.00
28 T	Bromochloromethane	0.372	0.341	8.3	96	0.00
29 T	Tetrahydrofuran	0.252	0.249	1.2	108	0.00
30 C	Chloroform	1.041	1.062	-2.0#	110	0.00
31 T	Cyclohexane	0.794	0.718	9.6	98	0.00
32 T	1,1,1-Trichloroethane	0.943	0.920	2.4	104	0.00
33 S	1,2-Dichloroethane-d4	0.652	0.578	11.3	95	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	99	0.00
35 S	Dibromofluoromethane	0.351	0.341	2.8	100	0.00
36 T	1,1-Dichloropropene	0.457	0.459	-0.4	105	0.00
37 T	Ethyl Acetate	0.464	0.494	-6.5	110	0.00
38 T	Carbon Tetrachloride	0.516	0.524	-1.6	102	0.00
39 T	Methylcyclohexane	0.577	0.532	7.8	96	0.00
40 TM	Benzene	1.339	1.408	-5.2	107	0.00
41 T	Methacrylonitrile	0.239	0.264	-10.5	115	0.00
42 TM	1,2-Dichloroethane	0.497	0.511	-2.8	106	0.00
43 T	Isopropyl Acetate	0.755	0.794	-5.2	110	0.00
44 TM	Trichloroethene	0.388	0.423	-9.0	113	0.00
45 C	1,2-Dichloropropane	0.328	0.353	-7.6#	112	0.00
46 T	Dibromomethane	0.257	0.279	-8.6	114	0.00
47 T	Bromodichloromethane	0.494	0.526	-6.5	108	0.00
48 T	Methyl methacrylate	0.347	0.374	-7.8	113	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX050322\
 Data File : VX028491.D
 Acq On : 03 May 2022 10:47
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: May 04 06:02:08 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X041922W.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 19 13:38:04 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	0.009	0.009	0.0	97	0.00
50 S	Toluene-d8	1.347	1.229	8.8	93	0.00
51 T	4-Methyl-2-Pentanone	0.485	0.519	-7.0	111	0.00
52 CM	Toluene	0.930	0.954	-2.6#	104	0.00
53 T	t-1,3-Dichloropropene	0.525	0.563	-7.2	105	0.00
54 T	cis-1,3-Dichloropropene	0.564	0.605	-7.3	108	0.00
55 T	1,1,2-Trichloroethane	0.378	0.422	-11.6	115	0.00
56 T	Ethyl methacrylate	0.548	0.625	-14.1	114	0.00
57 T	1,3-Dichloropropane	0.610	0.671	-10.0	114	0.00
58 T	2-Chloroethyl Vinyl ether	0.257	0.260	-1.2	97	0.00
59 T	2-Hexanone	0.375	0.401	-6.9	109	0.00
60 T	Dibromochloromethane	0.412	0.453	-10.0	108	0.00
61 T	1,2-Dibromoethane	0.415	0.454	-9.4	109	0.00
62 S	4-Bromofluorobenzene	0.515	0.482	6.4	93	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	95	0.00
64 T	Tetrachloroethene	0.391	0.438	-12.0	115	0.00
65 PM	Chlorobenzene	1.066	1.160	-8.8	108	0.00
66 T	1,1,1,2-Tetrachloroethane	0.397	0.454	-14.4	112	0.00
67 C	Ethyl Benzene	1.788	1.961	-9.7#	105	0.00
68 T	m/p-Xylenes	0.728	0.793	-8.9	103	0.00
69 T	o-Xylene	0.725	0.790	-9.0	105	0.00
70 T	Styrene	1.175	1.340	-14.0	107	0.00
71 P	Bromoform	0.313	0.366	-16.9	110	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	91	0.00
73 T	Isopropylbenzene	3.437	3.723	-8.3	106	0.00
74 T	N-amyl acetate	1.188	1.356	-14.1	111	0.00
75 P	1,1,2,2-Tetrachloroethane	1.115	1.266	-13.5	112	0.00
76 T	1,2,3-Trichloropropane	0.984	1.121	-13.9	109	0.00
77 T	Bromobenzene	0.895	0.977	-9.2	108	0.00
78 T	n-propylbenzene	3.811	4.221	-10.8	105	0.00
79 T	2-Chlorotoluene	2.369	2.523	-6.5	103	0.00
80 T	1,3,5-Trimethylbenzene	2.917	3.160	-8.3	102	0.00
81 T	trans-1,4-Dichloro-2-butene	0.331	0.369	-11.5	104	0.00
82 T	4-Chlorotoluene	2.729	2.945	-7.9	103	0.00
83 T	tert-Butylbenzene	2.979	3.196	-7.3	102	0.00
84 T	1,2,4-Trimethylbenzene	2.900	3.167	-9.2	101	0.00
85 T	sec-Butylbenzene	3.594	3.874	-7.8	101	0.00
86 T	p-Isopropyltoluene	3.100	3.327	-7.3	99	0.00
87 T	1,3-Dichlorobenzene	1.682	1.874	-11.4	105	0.00
88 T	1,4-Dichlorobenzene	1.766	1.899	-7.5	104	0.00
89 T	n-Butylbenzene	2.546	2.767	-8.7	100	0.00
90 T	Hexachloroethane	0.498	0.557	-11.8	99	0.00
91 T	1,2-Dichlorobenzene	1.671	1.866	-11.7	106	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.248	0.265	-6.9	103	0.00
93 T	1,2,4-Trichlorobenzene	0.994	1.181	-18.8	115	0.00
94 T	Hexachlorobutadiene	0.443	0.495	-11.7	112	0.00
95 T	Naphthalene	3.310	3.926	-18.6	109	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX050322\
Data File : VX028491.D
Acq On : 03 May 2022 10:47
Operator : JC/MD
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 2 Sample Multiplier: 1

Instrument :
MSVOA_X
LabSampleId :
VSTDCCC050

Quant Time: May 04 06:02:08 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X041922W.M
Quant Title : SW846 8260
QLast Update : Tue Apr 19 13:38:04 2022
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	0.975	1.164	-19.4	114	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 6