

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX050323\
 Data File : VX035509.D
 Acq On : 03 May 2023 15:01
 Operator : JC/MD
 Sample : 02533-15RE
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 EB-01

Quant Time: May 04 03:57:15 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X042523W.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 25 16:31:40 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	224971	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	365748	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	333325	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	153969	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	132684	48.886	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	97.780%
35) Dibromofluoromethane	5.385	113	111283	49.572	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	99.140%
50) Toluene-d8	8.647	98	426461	50.360	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	100.720%
62) 4-Bromofluorobenzene	11.079	95	162725	51.081	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	102.160%
Target Compounds						
					Qvalue	
16) Acetone	2.392	43	7519	6.379	ug/l	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX050323\
Data File : VX035509.D
Acq On : 03 May 2023 15:01
Operator : JC/MD
Sample : 02533-15RE
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
EB-01

Quant Time: May 04 03:57:15 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X042523W.M
Quant Title : SW846 8260
QLast Update : Tue Apr 25 16:31:40 2023
Response via : Initial Calibration

