

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX050323\
 Data File : VX035514.D
 Acq On : 03 May 2023 16:57
 Operator : JC/MD
 Sample : 02561-03
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 STORAGE-BLANK-WATER-REF-4

Quant Time: May 04 04:00:19 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X042523W.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 25 16:31:40 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	215608	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	350841	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	327557	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	156446	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	126019	48.447	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	96.900%
35) Dibromofluoromethane	5.385	113	105804	49.134	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	98.260%
50) Toluene-d8	8.647	98	420326	51.744	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	103.480%
62) 4-Bromofluorobenzene	11.079	95	162213	53.084	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	106.160%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX050323\
Data File : VX035514.D
Acq On : 03 May 2023 16:57
Operator : JC/MD
Sample : 02561-03
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
STORAGE-BLANK-WATER-REF-4

Quant Time: May 04 04:00:19 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X042523W.M
Quant Title : SW846 8260
QLast Update : Tue Apr 25 16:31:40 2023
Response via : Initial Calibration

