

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX050323\
 Data File : VX035519.D
 Acq On : 03 May 2023 18:53
 Operator : JC/MD
 Sample : 02597-02
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 25 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW-3-20230503

Quant Time: May 04 04:02:19 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X042523W.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 25 16:31:40 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	209961	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	334662	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	306361	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	142625	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	122596	48.398	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	96.800%
35) Dibromofluoromethane	5.385	113	104145	50.702	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	101.400%
50) Toluene-d8	8.647	98	392314	50.630	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	101.260%
62) 4-Bromofluorobenzene	11.079	95	148478	50.938	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	101.880%
Target Compounds						
16) Acetone	2.392	43	3807	3.460	ug/l	Qvalue 89

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX050323\
Data File : VX035519.D
Acq On : 03 May 2023 18:53
Operator : JC/MD
Sample : 02597-02
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 25 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW-3-20230503

Quant Time: May 04 04:02:19 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X042523W.M
Quant Title : SW846 8260
QLast Update : Tue Apr 25 16:31:40 2023
Response via : Initial Calibration

