

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX050323\
 Data File : VX035520.D
 Acq On : 03 May 2023 19:17
 Operator : JC/MD
 Sample : 02597-03
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 26 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW-2-20230503

Quant Time: May 04 04:02:44 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X042523W.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 25 16:31:40 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	201646	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	319969	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	293956	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	139118	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	119163	48.983	ug/l	0.00
Spiked Amount 50.000	Range 78 - 117		Recovery =	97.960%		
35) Dibromofluoromethane	5.391	113	99834	50.835	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery =	101.680%		
50) Toluene-d8	8.647	98	375225	50.649	ug/l	0.00
Spiked Amount 50.000	Range 92 - 112		Recovery =	101.300%		
62) 4-Bromofluorobenzene	11.079	95	143999	51.670	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery =	103.340%		
Target Compounds						
16) Acetone	2.386	43	1569	1.485	ug/l	Qvalue 100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX050323\
Data File : VX035520.D
Acq On : 03 May 2023 19:17
Operator : JC/MD
Sample : 02597-03
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 26 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW-2-20230503

Quant Time: May 04 04:02:44 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X042523W.M
Quant Title : SW846 8260
QLast Update : Tue Apr 25 16:31:40 2023
Response via : Initial Calibration

