

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX050323\
 Data File : VX035521.D
 Acq On : 03 May 2023 19:40
 Operator : JC/MD
 Sample : 02597-07
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 27 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW-1-20230503-A

Quant Time: May 04 04:03:10 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X042523W.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 25 16:31:40 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	213267	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	341230	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	316707	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	155913	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	125771	48.882	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	97.760%
35) Dibromofluoromethane	5.385	113	106517	50.859	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	101.720%
50) Toluene-d8	8.647	98	402925	50.999	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	102.000%
62) 4-Bromofluorobenzene	11.079	95	158271	53.253	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	106.500%
Target Compounds						
					Qvalue	
16) Acetone	2.386	43	34672	31.027	ug/l	98
25) 2-Butanone	4.581	43	10034	6.152	ug/l	93
40) Benzene	6.044	78	15454	1.647	ug/l	99
52) Toluene	8.720	92	2308	0.388	ug/l	99
68) m/p-Xylenes	10.305	106	1176	0.259	ug/l	93
95) Naphthalene	13.774	128	363324	34.868	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX050323\
Data File : VX035521.D
Acq On : 03 May 2023 19:40
Operator : JC/MD
Sample : 02597-07
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 27 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW-1-20230503-A

Quant Time: May 04 04:03:10 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X042523W.M
Quant Title : SW846 8260
QLast Update : Tue Apr 25 16:31:40 2023
Response via : Initial Calibration

