

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX050323\
 Data File : VX035522.D
 Acq On : 03 May 2023 20:03
 Operator : JC/MD
 Sample : 02597-04
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 28 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW-1-20230503

Manual Integrations
 APPROVED

Reviewed By :Krupa Patel 05/04/2023
 Supervised By :Mahesh Dadoda 05/04/2023

Quant Time: May 04 04:03:37 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X042523W.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 25 16:31:40 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	198477	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	318744	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	295103	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	144436	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	118430	49.459	ug/l	0.00
Spiked Amount 50.000	Range 78 - 117		Recovery = 98.920%			
35) Dibromofluoromethane	5.385	113	99536	50.878	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery = 101.760%			
50) Toluene-d8	8.647	98	373848	50.657	ug/l	0.00
Spiked Amount 50.000	Range 92 - 112		Recovery = 101.320%			
62) 4-Bromofluorobenzene	11.079	95	147160	53.007	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery = 106.020%			
Target Compounds						
16) Acetone	2.386	43	35275	33.919	ug/l	97
25) 2-Butanone	4.574	43	10846	7.145	ug/l	96
40) Benzene	6.044	78	14887	1.699	ug/l	94
95) Naphthalene	13.774	128	373998m	38.744	ug/l	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX050323\
Data File : VX035522.D
Acq On : 03 May 2023 20:03
Operator : JC/MD
Sample : 02597-04
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 28 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW-1-20230503

Quant Time: May 04 04:03:37 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X042523W.M
Quant Title : SW846 8260
QLast Update : Tue Apr 25 16:31:40 2023
Response via : Initial Calibration

Manual Integrations
APPROVED

Reviewed By :Krupa Patel 05/04/2023
Supervised By :Mahesh Dadoda 05/04/2023

