



Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX050323\
Data File : VX035495.D
Acq On : 03 May 2023 08:39
Operator : JC/MD
Sample : BFB
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X042523W.M
Title : SW846 8260
Last Update : Tue Apr 25 16:31:40 2023



AutoFind: Scans 1638, 1639, 1640; Background Corrected with Scan 1631

Target	Rel. to	Lower	Upper	Rel.	Raw	Result
Mass	Mass	Limit%	Limit%	Abn%	Abn	Pass/Fail
50	95	15	40	16.2	14169	PASS
75	95	30	60	49.1	42955	PASS
95	95	100	100	100.0	87568	PASS
96	95	5	9	6.7	5877	PASS
173	174	0.00	2	0.4	271	PASS
174	95	50	100	79.3	69427	PASS
175	174	5	9	7.7	5328	PASS
176	174	95	101	95.8	66485	PASS
177	176	5	9	6.9	4569	PASS