

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX050420\  
 Data File : VX016044.D  
 Acq On : 04 May 2020 18:39  
 Operator : JC/SP  
 Sample : L2469-16  
 Misc : 5.0mL/MSVOA X/WATER  
 ALS Vial : 19 Sample Multiplier: 1

Instrument :  
 MSVOA\_X  
 ClientSampleId :  
 SA-SW124-20200428

Quant Time: May 05 08:23:31 2020  
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA\_X\METHOD\82X050420W.M  
 Quant Title : SW846 8260  
 QLast Update : Tue May 05 07:11:41 2020  
 Response via : Initial Calibration

| Internal Standards          | R.T.  | QIon | Response | Conc  | Units   | Dev(Min)  |
|-----------------------------|-------|------|----------|-------|---------|-----------|
| 1) Pentafluorobenzene       | 5.64  | 168  | 297915   | 50.00 | ug/l    | 0.00      |
| 34) 1,4-Difluorobenzene     | 6.84  | 114  | 476667   | 50.00 | ug/l    | 0.00      |
| 63) Chlorobenzene-d5        | 10.10 | 117  | 465091   | 50.00 | ug/l    | 0.00      |
| 72) 1,4-Dichlorobenzene-d4  | 12.06 | 152  | 241835   | 50.00 | ug/l    | 0.00      |
| System Monitoring Compounds |       |      |          |       |         |           |
| 33) 1,2-Dichloroethane-d4   | 6.03  | 65   | 181376   | 46.27 | ug/l    | 0.00      |
| Spiked Amount 50.000        |       |      | Recovery | =     | 92.54%  |           |
| 35) Dibromofluoromethane    | 5.47  | 113  | 137134   | 45.41 | ug/l    | 0.00      |
| Spiked Amount 50.000        |       |      | Recovery | =     | 90.82%  |           |
| 50) Toluene-d8              | 8.70  | 98   | 584997   | 50.28 | ug/l    | 0.00      |
| Spiked Amount 50.000        |       |      | Recovery | =     | 100.56% |           |
| 62) 4-Bromofluorobenzene    | 11.12 | 95   | 239329   | 54.08 | ug/l    | 0.00      |
| Spiked Amount 50.000        |       |      | Recovery | =     | 108.16% |           |
| Target Compounds            |       |      |          |       |         |           |
| 16) Acetone                 | 2.42  | 43   | 4517     | 2.582 | ug/l    | Qvalue 90 |

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX050420\  
Data File : VX016044.D  
Acq On : 04 May 2020 18:39  
Operator : JC/SP  
Sample : L2469-16  
Misc : 5.0mL/MSVOA X/WATER  
ALS Vial : 19 Sample Multiplier: 1

Instrument :  
MSVOA\_X  
ClientSampleId :  
SA-SW124-20200428

Quant Time: May 05 08:23:31 2020  
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA\_X\METHOD\82X050420W.M  
Quant Title : SW846 8260  
QLast Update : Tue May 05 07:11:41 2020  
Response via : Initial Calibration

