

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX050420\
 Data File : VX016047.D
 Acq On : 04 May 2020 19:47
 Operator : JC/SP
 Sample : L2469-19
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 SA-PZ124-20200428

Quant Time: May 05 09:02:09 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X050420W.M
 Quant Title : SW846 8260
 QLast Update : Tue May 05 07:11:41 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.64	168	305925	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.84	114	500399	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	482125	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	249790	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.03	65	189439	47.06	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.12%	
35) Dibromofluoromethane	5.47	113	146719	46.28	ug/l	0.00
Spiked Amount 50.000			Recovery	=	92.56%	
50) Toluene-d8	8.70	98	607767	49.76	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.52%	
62) 4-Bromofluorobenzene	11.12	95	242571	52.21	ug/l	0.00
Spiked Amount 50.000			Recovery	=	104.42%	
Target Compounds						
6) Chloroethane	1.72	64	5651	2.315	ug/l	98
12) 1,1-Dichloroethene	2.36	96	3465	1.090	ug/l	84
16) Acetone	2.42	43	3484	1.940	ug/l	90
24) 1,1-Dichloroethane	3.67	63	42656	6.932	ug/l	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX050420\
Data File : VX016047.D
Acq On : 04 May 2020 19:47
Operator : JC/SP
Sample : L2469-19
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
SA-PZ124-20200428

Quant Time: May 05 09:02:09 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X050420W.M
Quant Title : SW846 8260
QLast Update : Tue May 05 07:11:41 2020
Response via : Initial Calibration

