

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX050422\
 Data File : VX028522.D
 Acq On : 04 May 2022 15:29
 Operator : JC/MD
 Sample : N2695-01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 EFF0522

Quant Time: May 05 03:18:43 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X041922W.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 19 13:38:04 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	410122	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	694273	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	649947	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	290592	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	275737	51.564	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	= 103.120%	
35) Dibromofluoromethane	5.391	113	259610	53.320	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	= 106.640%	
50) Toluene-d8	8.653	98	925922	49.513	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	= 99.020%	
62) 4-Bromofluorobenzene	11.079	95	324718	45.413	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	= 90.820%	
Target Compounds						
11) Tert butyl alcohol	2.953	59	22350	12.382	ug/l	Qvalue 97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX050422\
Data File : VX028522.D
Acq On : 04 May 2022 15:29
Operator : JC/MD
Sample : N2695-01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
EFF0522

Quant Time: May 05 03:18:43 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X041922W.M
Quant Title : SW846 8260
QLast Update : Tue Apr 19 13:38:04 2022
Response via : Initial Calibration

