

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_X\Data\VX050422\  
 Data File : VX028526.D  
 Acq On : 04 May 2022 17:26  
 Operator : JC/MD  
 Sample : VSTDCCC050  
 Misc : 5.0mL/MSVOA\_X/WATER  
 ALS Vial : 11 Sample Multiplier: 1

Instrument :  
 MSVOA\_X  
 ClientSampleId :  
 VSTDCCC050EC

Manual Integrations  
 APPROVED

Reviewed By : John Carlone 05/05/2022  
 Supervised By : Mahesh Dadoda 05/05/2022

Quant Time: May 05 03:19:45 2022  
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_X\Method\82X041922W.M  
 Quant Title : SW846 8260  
 QLast Update : Tue Apr 19 13:38:04 2022  
 Response via : Initial Calibration

| Compound                     | R.T.           | QIon | Response   | Conc     | Units  | Dev(Min) |
|------------------------------|----------------|------|------------|----------|--------|----------|
| -----                        |                |      |            |          |        |          |
| Internal Standards           |                |      |            |          |        |          |
| 1) Pentafluorobenzene        | 5.550          | 168  | 347027     | 50.000   | ug/l   | 0.00     |
| 34) 1,4-Difluorobenzene      | 6.763          | 114  | 534796     | 50.000   | ug/l   | 0.00     |
| 63) Chlorobenzene-d5         | 10.055         | 117  | 505006     | 50.000   | ug/l   | 0.00     |
| 72) 1,4-Dichlorobenzene-d4   | 12.024         | 152  | 274575     | 50.000   | ug/l   | 0.00     |
|                              |                |      |            |          |        |          |
| System Monitoring Compounds  |                |      |            |          |        |          |
| 33) 1,2-Dichloroethane-d4    | 5.952          | 65   | 207689     | 45.900   | ug/l   | 0.00     |
| Spiked Amount 50.000         | Range 61 - 141 |      | Recovery = | 91.800%  |        |          |
| 35) Dibromofluoromethane     | 5.385          | 113  | 193976     | 51.720   | ug/l   | 0.00     |
| Spiked Amount 50.000         | Range 69 - 133 |      | Recovery = | 103.440% |        |          |
| 50) Toluene-d8               | 8.647          | 98   | 689516     | 47.867   | ug/l   | 0.00     |
| Spiked Amount 50.000         | Range 65 - 126 |      | Recovery = | 95.740%  |        |          |
| 62) 4-Bromofluorobenzene     | 11.079         | 95   | 270059     | 49.032   | ug/l   | 0.00     |
| Spiked Amount 50.000         | Range 58 - 135 |      | Recovery = | 98.060%  |        |          |
|                              |                |      |            |          |        |          |
| Target Compounds             |                |      |            |          | Qvalue |          |
| 2) Dichlorodifluoromethane   | 1.166          | 85   | 131603     | 58.748   | ug/l   | 100      |
| 3) Chloromethane             | 1.288          | 50   | 128852     | 57.413   | ug/l   | 99       |
| 4) Vinyl Chloride            | 1.374          | 62   | 158018     | 45.168   | ug/l   | 99       |
| 5) Bromomethane              | 1.599          | 94   | 154179     | 49.360   | ug/l   | 99       |
| 6) Chloroethane              | 1.679          | 64   | 113787     | 45.107   | ug/l   | 99       |
| 7) Trichlorofluoromethane    | 1.880          | 101  | 300442     | 38.945   | ug/l   | 98       |
| 8) Diethyl Ether             | 2.130          | 74   | 113653     | 48.970   | ug/l   | 95       |
| 9) 1,1,2-Trichlorotrifluo... | 2.325          | 101  | 166401     | 59.213   | ug/l   | 97       |
| 10) Methyl Iodide            | 2.453          | 142  | 189894     | 49.472   | ug/l   | 98       |
| 11) Tert butyl alcohol       | 2.971          | 59   | 191264     | 219.511  | ug/l   | 99       |
| 12) 1,1-Dichloroethene       | 2.319          | 96   | 146650     | 54.987   | ug/l   | 92       |
| 13) Acrolein                 | 2.233          | 56   | 106984     | 192.923  | ug/l   | 100      |
| 14) Allyl chloride           | 2.660          | 41   | 214537     | 47.940   | ug/l   | 94       |
| 15) Acrylonitrile            | 3.062          | 53   | 468192     | 253.907  | ug/l   | 99       |
| 16) Acetone                  | 2.380          | 43   | 381773     | 276.389  | ug/l   | 96       |
| 17) Carbon Disulfide         | 2.508          | 76   | 307461     | 42.810   | ug/l   | 99       |
| 18) Methyl Acetate           | 2.703          | 43   | 205303     | 51.031   | ug/l   | 99       |
| 19) Methyl tert-butyl Ether  | 3.111          | 73   | 571804     | 51.415   | ug/l   | 99       |
| 20) Methylene Chloride       | 2.788          | 84   | 178741     | 53.239   | ug/l   | 99       |
| 21) trans-1,2-Dichloroethene | 3.093          | 96   | 167717     | 46.241   | ug/l   | 91       |
| 22) Diisopropyl ether        | 3.757          | 45   | 497987     | 48.857   | ug/l # | 87       |
| 23) Vinyl Acetate            | 3.721          | 43   | 2150758    | 241.528  | ug/l   | 98       |
| 24) 1,1-Dichloroethane       | 3.605          | 63   | 301566     | 50.107   | ug/l   | 99       |
| 25) 2-Butanone               | 4.550          | 43   | 629042     | 234.737  | ug/l   | 99       |
| 26) 2,2-Dichloropropane      | 4.477          | 77   | 255568     | 43.492   | ug/l   | 98       |
| 27) cis-1,2-Dichloroethene   | 4.489          | 96   | 212570     | 49.152   | ug/l   | 92       |
| 28) Bromochloromethane       | 4.897          | 49   | 112575     | 43.598   | ug/l   | 88       |
| 29) Tetrahydrofuran          | 5.001          | 42   | 395153     | 226.149  | ug/l   | 99       |
| 30) Chloroform               | 5.093          | 83   | 348580     | 48.228   | ug/l   | 98       |
| 31) Cyclohexane              | 5.471          | 56   | 233032     | 42.271   | ug/l   | 95       |
| 32) 1,1,1-Trichloroethane    | 5.385          | 97   | 303315     | 46.344   | ug/l   | 99       |
| 36) 1,1-Dichloropropene      | 5.690          | 75   | 233359     | 47.701   | ug/l   | 98       |
| 37) Ethyl Acetate            | 4.708          | 43   | 245864     | 49.529   | ug/l   | 99       |
| 38) Carbon Tetrachloride     | 5.678          | 117  | 268552     | 48.627   | ug/l   | 99       |
| 39) Methylcyclohexane        | 7.379          | 83   | 270613     | 43.840   | ug/l   | 95       |
| 40) Benzene                  | 6.037          | 78   | 707652     | 49.416   | ug/l   | 100      |

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_X\Data\VX050422\  
 Data File : VX028526.D  
 Acq On : 04 May 2022 17:26  
 Operator : JC/MD  
 Sample : VSTDCCC050  
 Misc : 5.0mL/MSVOA\_X/WATER  
 ALS Vial : 11 Sample Multiplier: 1

Instrument :  
 MSVOA\_X  
 ClientSampleId :  
 VSTDCCC050EC

Manual Integrations  
 APPROVED

Reviewed By : John Carlone 05/05/2022  
 Supervised By : Mahesh Dadoda 05/05/2022

Quant Time: May 05 03:19:45 2022  
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_X\Method\82X041922W.M  
 Quant Title : SW846 8260  
 QLast Update : Tue Apr 19 13:38:04 2022  
 Response via : Initial Calibration

| Compound                      | R.T.   | QIon | Response | Conc    | Units  | Dev(Min) |
|-------------------------------|--------|------|----------|---------|--------|----------|
| 41) Methacrylonitrile         | 4.916  | 41   | 132176   | 51.609  | ug/l   | 99       |
| 42) 1,2-Dichloroethane        | 6.086  | 62   | 253993   | 47.781  | ug/l   | 100      |
| 43) Isopropyl Acetate         | 6.336  | 43   | 396742   | 49.108  | ug/l   | 98       |
| 44) Trichloroethene           | 7.123  | 130  | 211529   | 50.937  | ug/l   | 100      |
| 45) 1,2-Dichloropropane       | 7.428  | 63   | 178095   | 50.808  | ug/l   | 97       |
| 46) Dibromomethane            | 7.580  | 93   | 138668   | 50.431  | ug/l   | 99       |
| 47) Bromodichloromethane      | 7.824  | 83   | 270139   | 51.116  | ug/l   | 98       |
| 48) Methyl methacrylate       | 7.690  | 41   | 185606   | 50.040  | ug/l   | 93       |
| 49) 1,4-Dioxane               | 7.659  | 88   | 87104    | 877.828 | ug/l   | 94       |
| 51) 4-Methyl-2-Pentanone      | 8.574  | 43   | 1300709  | 250.891 | ug/l   | 98       |
| 52) Toluene                   | 8.720  | 92   | 483506   | 48.599  | ug/l   | 99       |
| 53) t-1,3-Dichloropropene     | 8.976  | 75   | 279409   | 49.747  | ug/l   | 99       |
| 54) cis-1,3-Dichloropropene   | 8.366  | 75   | 304083   | 50.417  | ug/l   | 95       |
| 55) 1,1,2-Trichloroethane     | 9.153  | 97   | 212420   | 52.549  | ug/l   | 98       |
| 56) Ethyl methacrylate        | 9.116  | 69   | 312297   | 53.297  | ug/l # | 94       |
| 57) 1,3-Dichloropropane       | 9.311  | 76   | 332652   | 50.986  | ug/l   | 99       |
| 58) 2-Chloroethyl Vinyl ether | 8.244  | 63   | 701030   | 254.987 | ug/l   | 94       |
| 59) 2-Hexanone                | 9.433  | 43   | 999626   | 248.918 | ug/l   | 96       |
| 60) Dibromochloromethane      | 9.525  | 129  | 231528   | 52.539  | ug/l   | 98       |
| 61) 1,2-Dibromoethane         | 9.610  | 107  | 225057   | 50.749  | ug/l   | 99       |
| 64) Tetrachloroethene         | 9.275  | 164  | 204001   | 51.717  | ug/l   | 98       |
| 65) Chlorobenzene             | 10.079 | 112  | 557431   | 51.771  | ug/l   | 100      |
| 66) 1,1,1,2-Tetrachloroethane | 10.165 | 131  | 217112   | 54.130  | ug/l   | 99       |
| 67) Ethyl Benzene             | 10.195 | 91   | 947300   | 52.469  | ug/l   | 96       |
| 68) m/p-Xylenes               | 10.305 | 106  | 763061   | 103.845 | ug/l   | 95       |
| 69) o-Xylene                  | 10.640 | 106  | 377721   | 51.558  | ug/l   | 97       |
| 70) Styrene                   | 10.659 | 104  | 641612   | 54.058  | ug/l   | 99       |
| 71) Bromoform                 | 10.799 | 173  | 182424   | 49.339  | ug/l   | 99       |
| 73) Isopropylbenzene          | 10.963 | 105  | 994743   | 52.709  | ug/l   | 99       |
| 74) N-amyl acetate            | 10.842 | 43   | 357533   | 54.811  | ug/l   | 96       |
| 75) 1,1,2,2-Tetrachloroethane | 11.213 | 83   | 335315   | 54.752  | ug/l   | 99       |
| 76) 1,2,3-Trichloropropane    | 11.238 | 75   | 291875m  | 54.013  | ug/l   |          |
| 77) Bromobenzene              | 11.201 | 156  | 260971   | 53.096  | ug/l   | 90       |
| 78) n-propylbenzene           | 11.305 | 91   | 1120860  | 53.560  | ug/l   | 97       |
| 79) 2-Chlorotoluene           | 11.366 | 91   | 674000   | 51.804  | ug/l   | 96       |
| 80) 1,3,5-Trimethylbenzene    | 11.451 | 105  | 834093   | 52.066  | ug/l   | 97       |
| 81) trans-1,4-Dichloro-2-b... | 11.018 | 75   | 93086    | 51.191  | ug/l   | 91       |
| 82) 4-Chlorotoluene           | 11.457 | 91   | 778214   | 51.926  | ug/l   | 96       |
| 83) tert-Butylbenzene         | 11.719 | 119  | 862232   | 52.708  | ug/l   | 99       |
| 84) 1,2,4-Trimethylbenzene    | 11.750 | 105  | 839677   | 52.728  | ug/l   | 98       |
| 85) sec-Butylbenzene          | 11.890 | 105  | 1036898  | 52.539  | ug/l   | 98       |
| 86) p-Isopropyltoluene        | 12.012 | 119  | 887161   | 52.106  | ug/l   | 98       |
| 87) 1,3-Dichlorobenzene       | 11.969 | 146  | 493221   | 53.405  | ug/l   | 99       |
| 88) 1,4-Dichlorobenzene       | 12.043 | 146  | 498863   | 51.445  | ug/l   | 99       |
| 89) n-Butylbenzene            | 12.335 | 91   | 736952   | 52.713  | ug/l   | 97       |
| 90) Hexachloroethane          | 12.542 | 117  | 151236   | 49.389  | ug/l   | 98       |
| 91) 1,2-Dichlorobenzene       | 12.335 | 146  | 490917   | 53.509  | ug/l   | 98       |
| 92) 1,2-Dibromo-3-Chloropr... | 12.945 | 75   | 70663    | 51.882  | ug/l   | 89       |
| 93) 1,2,4-Trichlorobenzene    | 13.591 | 180  | 310378   | 56.880  | ug/l   | 100      |
| 94) Hexachlorobutadiene       | 13.725 | 225  | 129191   | 53.057  | ug/l   | 95       |
| 95) Naphthalene               | 13.774 | 128  | 1034655  | 56.919  | ug/l   | 100      |
| 96) 1,2,3-Trichlorobenzene    | 13.963 | 180  | 306378   | 57.201  | ug/l   | 97       |

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_X\Data\VX050422\  
Data File : VX028526.D  
Acq On : 04 May 2022 17:26  
Operator : JC/MD  
Sample : VSTDCCC050  
Misc : 5.0mL/MSVOA\_X/WATER  
ALS Vial : 11 Sample Multiplier: 1

Instrument :  
MSVOA\_X  
ClientSampleId :  
VSTDCCC050EC

Manual Integrations  
APPROVED

Reviewed By :John Carlone 05/05/2022  
Supervised By :Mahesh Dadoda 05/05/2022

Quant Time: May 05 03:19:45 2022  
Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_X\Method\82X041922W.M  
Quant Title : SW846 8260  
QLast Update : Tue Apr 19 13:38:04 2022  
Response via : Initial Calibration

| Compound | R.T. | QIon | Response | Conc | Units | Dev(Min) |
|----------|------|------|----------|------|-------|----------|
|----------|------|------|----------|------|-------|----------|

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Data Path : Z:\voasrv\HPCHEM1\MSVOA\_X\Data\VX050422\  
Data File : VX028526.D  
Acq On : 04 May 2022 17:26  
Operator : JC/MD  
Sample : VSTDCCC050  
Misc : 5.0mL/MSVOA\_X/WATER  
ALS Vial : 11 Sample Multiplier: 1


Instrument :  
MSVOA\_X  
ClientSampleId :  
VSTDCCC050EC

Quant Time: May 05 03:19:45 2022  
Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_X\Method\82X041922W.M  
Quant Title : SW846 8260  
QLast Update : Tue Apr 19 13:38:04 2022  
Response via : Initial Calibration

Manual Integrations  
APPROVED


Reviewed By : John Carlone 05/05/2022  
Supervised By : Mahesh Dadoda 05/05/2022





#1  
Pentafluorobenzene  
Concen: 50.000 ug/l  
RT: 5.550 min Scan# 71  
Delta R.T. -0.006 min  
Lab File: VX028526.D  
Acq: 04 May 2022 17:26



Instrument :  
MSVOA\_X  
ClientSampleId :  
VSTDCCC050EC



Tgt Ion:168 Resp: 34702  
Ion Ratio Lower Upper  
168 100  
99 53.6 41.7 62.5




Manual Integrations  
APPROVED


Reviewed By :John Carlone 05/05/2022  
Supervised By :Mahesh Dadoda 05/05/2022



#2  
Dichlorodifluoromethane  
Concen: 58.748 ug/l  
RT: 1.166 min Scan# 13  
Delta R.T. 0.000 min  
Lab File: VX028526.D  
Acq: 04 May 2022 17:26


Tgt Ion: 85 Resp: 131603  
Ion Ratio Lower Upper  
85 100  
87 32.4 16.2 48.6





#3  
Chloromethane  
Concen: 57.413 ug/l  
RT: 1.288 min Scan# 31  
Delta R.T. -0.006 min  
Lab File: VX028526.D  
Acq: 04 May 2022 17:26




Instrument :  
MSVOA\_X  
ClientSampleId :  
VSTDCCC050EC



Tgt Ion: 50 Resp: 128852  
Ion Ratio Lower Upper  
50 100  
52 32.5 26.4 39.6




Manual Integrations  
APPROVED


Reviewed By :John Carlone 05/05/2022  
Supervised By :Mahesh Dadoda 05/05/2022



#4  
Vinyl Chloride  
Concen: 45.168 ug/l  
RT: 1.374 min Scan# 47  
Delta R.T. -0.000 min  
Lab File: VX028526.D  
Acq: 04 May 2022 17:26


Tgt Ion: 62 Resp: 158018  
Ion Ratio Lower Upper  
62 100  
64 31.3 25.5 38.3





#5  
Bromomethane  
Concen: 49.360 ug/l  
RT: 1.599 min Scan# 84  
Delta R.T. 0.000 min  
Lab File: VX028526.D  
Acq: 04 May 2022 17:26




Instrument :  
MSVOA\_X  
ClientSampleId :  
VSTDCCC050EC




Tgt Ion: 94 Resp: 154179  
Ion Ratio Lower Upper  
94 100  
96 95.2 75.3 112.9

Manual Integrations  
APPROVED



Reviewed By :John Carlone 05/05/2022  
Supervised By :Mahesh Dadoda 05/05/2022




#6  
Chloroethane  
Concen: 45.107 ug/l  
RT: 1.679 min Scan# 97  
Delta R.T. 0.001 min  
Lab File: VX028526.D  
Acq: 04 May 2022 17:26




Tgt Ion: 64 Resp: 113787  
Ion Ratio Lower Upper  
64 100  
66 32.0 25.3 37.9





#7  
 Trichlorofluoromethane  
 Concen: 38.945 ug/l  
 RT: 1.880 min Scan# 11  
 Delta R.T. -0.006 min  
 Lab File: VX028526.D  
 Acq: 04 May 2022 17:26




Instrument :  
 MSVOA\_X  
 ClientSampleId :  
 VSTDCCC050EC



Tgt Ion:101 Resp: 30044  
 Ion Ratio Lower Upper  
 101 100  
 103 64.3 50.5 75.7




Manual Integrations  
 APPROVED

Reviewed By :John Carlone 05/05/2022  
 Supervised By :Mahesh Dadoda 05/05/2022




#8  
 Diethyl Ether  
 Concen: 48.970 ug/l  
 RT: 2.130 min Scan# 171  
 Delta R.T. -0.000 min  
 Lab File: VX028526.D  
 Acq: 04 May 2022 17:26

Tgt Ion: 74 Resp: 113653  
 Ion Ratio Lower Upper  
 74 100  
 45 79.6 42.0 126.0











#9  
1,1,2-Trichlorotrifluoroethane  
Concen: 59.213 ug/l  
RT: 2.325 min Scan# 204  
Delta R.T. -0.006 min  
Lab File: VX028526.D  
Acq: 04 May 2022 17:26

Instrument :  
MSVOA\_X  
ClientSampleId :  
VSTDCCC050EC

Tgt Ion:101 Resp: 16640  
Ion Ratio Lower Upper  
101 100  
85 43.6 33.5 50.3  
151 79.4 61.8 92.6




Manual Integrations  
APPROVED


Reviewed By :John Carlone 05/05/2022  
Supervised By :Mahesh Dadoda 05/05/2022



#10  
Methyl Iodide  
Concen: 49.472 ug/l  
RT: 2.453 min Scan# 224  
Delta R.T. -0.000 min  
Lab File: VX028526.D  
Acq: 04 May 2022 17:26

Tgt Ion:142 Resp: 189894  
Ion Ratio Lower Upper  
142 100  
127 44.2 33.9 50.9  
141 13.9 11.4 17.2





#11

Tert butyl alcohol

Concen: 219.511 ug/l

RT: 2.971 min Scan# 308

Delta R.T. 0.006 min

Lab File: VX028526.D

Acq: 04 May 2022 17:26

Instrument :

MSVOA\_X

ClientSampleId :

VSTDCCC050EC

Tgt Ion: 59 Resp: 191264

Ion Ratio Lower Upper

59 100





57 9.9 8.3 12.5

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/05/2022

Supervised By :Mahesh Dadoda 05/05/2022



#12

1,1-Dichloroethene

Concen: 54.987 ug/l

RT: 2.319 min Scan# 202

Delta R.T. -0.000 min

Lab File: VX028526.D

Acq: 04 May 2022 17:26




Tgt Ion: 96 Resp: 146650


Ion Ratio Lower Upper

96 100

61 142.2 124.6 187.0


98 63.2 51.0 76.4





#13  
Acrolein  
Concen: 192.923 ug/l  
RT: 2.233 min Scan# 11  
Delta R.T. 0.000 min  
Lab File: VX028526.D  
Acq: 04 May 2022 17:26




Instrument :  
MSVOA\_X  
ClientSampleId :  
VSTDCCC050EC



Tgt Ion: 56 Resp: 106984  
Ion Ratio Lower Upper  
56 100  
55 69.0 55.0 82.6




Manual Integrations  
APPROVED


Reviewed By :John Carlone 05/05/2022  
Supervised By :Mahesh Dadoda 05/05/2022



#14  
Allyl chloride  
Concen: 47.940 ug/l  
RT: 2.660 min Scan# 258  
Delta R.T. -0.006 min  
Lab File: VX028526.D  
Acq: 04 May 2022 17:26


Tgt Ion: 41 Resp: 214537  
Ion Ratio Lower Upper  
41 100  
39 70.7 50.4 75.6  
76 41.5 32.7 49.1





#15  
Acrylonitrile  
Concen: 253.907 ug/l  
RT: 3.062 min Scan# 31  
Delta R.T. 0.000 min  
Lab File: VX028526.D  
Acq: 04 May 2022 17:26




Instrument :  
MSVOA\_X  
ClientSampleId :  
VSTDCCC050EC




Tgt Ion: 53 Resp: 468192  
Ion Ratio Lower Upper  
53 100  
52 82.7 65.7 98.5  
51 36.0 28.4 42.6

Manual Integrations  
APPROVED



Reviewed By :John Carlone 05/05/2022  
Supervised By :Mahesh Dadoda 05/05/2022




#16  
Acetone  
Concen: 276.389 ug/l  
RT: 2.380 min Scan# 212  
Delta R.T. -0.000 min  
Lab File: VX028526.D  
Acq: 04 May 2022 17:26



Tgt Ion: 43 Resp: 381773  
Ion Ratio Lower Upper  
43 100  
58 33.7 28.8 43.2






#17  
Carbon Disulfide  
Concen: 42.810 ug/l  
RT: 2.508 min Scan# 21  
Delta R.T. -0.006 min  
Lab File: VX028526.D  
Acq: 04 May 2022 17:26

Instrument :

MSVOA\_X

ClientSampleId :

VSTDCCC050EC






Tgt Ion: 76 Resp: 30746  
Ion Ratio Lower Upper  
76 100  
78 9.2 7.2 10.8

Manual Integrations




APPROVED


Reviewed By :John Carlone 05/05/2022  
Supervised By :Mahesh Dadoda 05/05/2022



#18  
Methyl Acetate  
Concen: 51.031 ug/l  
RT: 2.703 min Scan# 265  
Delta R.T. -0.000 min  
Lab File: VX028526.D  
Acq: 04 May 2022 17:26

Tgt Ion: 43 Resp: 205303  
Ion Ratio Lower Upper  
43 100  
74 29.1 22.8 34.2









#19  
Methyl tert-butyl Ether  
Concen: 51.415 ug/l  
RT: 3.111 min Scan# 311  
Delta R.T. 0.000 min  
Lab File: VX028526.D  
Acq: 04 May 2022 17:26

Instrument :  
MSVOA\_X  
ClientSampleId :  
VSTDCCC050EC

Tgt Ion: 73 Resp: 571804  
Ion Ratio Lower Upper  
73 100  
57 20.5 16.8 25.2




Manual Integrations  
APPROVED


Reviewed By :John Carlone 05/05/2022  
Supervised By :Mahesh Dadoda 05/05/2022



#20  
Methylene Chloride  
Concen: 53.239 ug/l  
RT: 2.788 min Scan# 279  
Delta R.T. 0.000 min  
Lab File: VX028526.D  
Acq: 04 May 2022 17:26


Tgt Ion: 84 Resp: 178741  
Ion Ratio Lower Upper  
84 100  
49 106.4 85.5 128.3  
51 32.4 26.6 39.8  
86 65.0 51.5 77.3





#21  
trans-1,2-Dichloroethene  
Concen: 46.241 ug/l  
RT: 3.093 min Scan# 31  
Delta R.T. -0.000 min  
Lab File: VX028526.D  
Acq: 04 May 2022 17:26




Instrument :  
MSVOA\_X  
ClientSampleId :  
VSTDCCC050EC




Tgt Ion: 96 Resp: 16771  
Ion Ratio Lower Upper  
96 100  
61 117.3 105.4 158.2  
98 64.7 50.0 75.0

Manual Integrations  
APPROVED



Reviewed By :John Carlone 05/05/2022  
Supervised By :Mahesh Dadoda 05/05/2022




#22  
Diisopropyl ether  
Concen: 48.857 ug/l  
RT: 3.757 min Scan# 438  
Delta R.T. -0.006 min  
Lab File: VX028526.D  
Acq: 04 May 2022 17:26



Tgt Ion: 45 Resp: 497987  
Ion Ratio Lower Upper  
45 100  
43 79.7 50.8 76.2#  
87 35.7 26.6 39.8  
59 12.9 10.5 15.7





#23  
 Vinyl Acetate  
 Concen: 241.528 ug/l  
 RT: 3.721 min Scan# 413  
 Delta R.T. -0.000 min  
 Lab File: VX028526.D  
 Acq: 04 May 2022 17:26

Instrument :

MSVOA\_X

ClientSampleId :





VSTDCCC050EC

Tgt Ion: 43 Resp: 2150758  
 Ion Ratio Lower Upper  
 43 100  
 86 14.2 10.8 16.2

Manual Integrations




APPROVED

Reviewed By :John Carlone 05/05/2022  
 Supervised By :Mahesh Dadoda 05/05/2022




#24  
 1,1-Dichloroethane  
 Concen: 50.107 ug/l  
 RT: 3.605 min Scan# 413  
 Delta R.T. -0.006 min  
 Lab File: VX028526.D  
 Acq: 04 May 2022 17:26

Tgt Ion: 63 Resp: 301566  
 Ion Ratio Lower Upper  
 63 100  
 98 8.0 3.9 11.7  
 100 5.2 2.4 7.2








#25  
2-Butanone  
Concen: 234.737 ug/l  
RT: 4.550 min Scan# 569  
Delta R.T. -0.006 min  
Lab File: VX028526.D  
Acq: 04 May 2022 17:26




Instrument :  
MSVOA\_X  
ClientSampleId :  
VSTDCCC050EC



Tgt Ion: 43 Resp: 62904  
Ion Ratio Lower Upper  
43 100  
72 31.1 24.4 36.6




Manual Integrations  
APPROVED


Reviewed By :John Carlone 05/05/2022  
Supervised By :Mahesh Dadoda 05/05/2022



#26  
2,2-Dichloropropane  
Concen: 43.492 ug/l  
RT: 4.477 min Scan# 556  
Delta R.T. -0.000 min  
Lab File: VX028526.D  
Acq: 04 May 2022 17:26

Tgt Ion: 77 Resp: 255568  
Ion Ratio Lower Upper  
77 100  
97 24.4 11.8 35.4






#27  
 cis-1,2-Dichloroethene  
 Concen: 49.152 ug/l  
 RT: 4.489 min Scan# 511  
 Delta R.T. 0.000 min  
 Lab File: VX028526.D  
 Acq: 04 May 2022 17:26

Instrument :

MSVOA\_X

ClientSampleId :

VSTDCCC050EC






Tgt Ion: 96 Resp: 212570  
 Ion Ratio Lower Upper  
 96 100  
 61 127.7 0.0 283.8  
 98 65.0 0.0 129.8

Manual Integrations

APPROVED




Reviewed By :John Carlone 05/05/2022


Supervised By :Mahesh Dadoda 05/05/2022



#28  
 Bromochloromethane  
 Concen: 43.598 ug/l  
 RT: 4.897 min Scan# 625  
 Delta R.T. 0.000 min  
 Lab File: VX028526.D  
 Acq: 04 May 2022 17:26


Tgt Ion: 49 Resp: 112575  
 Ion Ratio Lower Upper  
 49 100  
 129 2.6 0.0 4.4  
 130 99.6 70.7 106.1





#29  
Tetrahydrofuran  
Concen: 226.149 ug/l  
RT: 5.001 min Scan# 642  
Delta R.T. -0.006 min  
Lab File: VX028526.D  
Acq: 04 May 2022 17:26




Instrument :  
MSVOA\_X  
ClientSampleId :  
VSTDCCC050EC



Tgt Ion: 42 Resp: 39515  
Ion Ratio Lower Upper  
42 100  
72 55.9 44.1 66.1  
71 51.3 40.5 60.7




Manual Integrations  
APPROVED


Reviewed By :John Carlone 05/05/2022  
Supervised By :Mahesh Dadoda 05/05/2022



#30  
Chloroform  
Concen: 48.228 ug/l  
RT: 5.093 min Scan# 657  
Delta R.T. -0.006 min  
Lab File: VX028526.D  
Acq: 04 May 2022 17:26


Tgt Ion: 83 Resp: 348580  
Ion Ratio Lower Upper  
83 100  
85 64.3 52.7 79.1





#31  
Cyclohexane  
Concen: 42.271 ug/l  
RT: 5.471 min Scan# 719  
Delta R.T. 0.001 min  
Lab File: VX028526.D  
Acq: 04 May 2022 17:26




Instrument :  
MSVOA\_X  
ClientSampleId :  
VSTDCCC050EC




Tgt Ion: 56 Resp: 23303  
Ion Ratio Lower Upper  
56 100  
69 33.2 26.7 40.1  
84 102.5 77.2 115.8

Manual Integrations  
APPROVED



Reviewed By :John Carlone 05/05/2022  
Supervised By :Mahesh Dadoda 05/05/2022




#32  
1,1,1-Trichloroethane  
Concen: 46.344 ug/l  
RT: 5.385 min Scan# 705  
Delta R.T. 0.000 min  
Lab File: VX028526.D  
Acq: 04 May 2022 17:26



Tgt Ion: 97 Resp: 303315  
Ion Ratio Lower Upper  
97 100  
99 64.6 51.8 77.6  
61 40.5 34.0 51.0





#33  
1,2-Dichloroethane-d4  
Concen: 45.900 ug/l  
RT: 5.952 min Scan# 799  
Delta R.T. -0.006 min  
Lab File: VX028526.D  
Acq: 04 May 2022 17:26

Instrument :

MSVOA\_X

ClientSampleId :





VSTDCCC050EC

Tgt Ion: 65 Resp: 207689  
Ion Ratio Lower Upper  
65 100  
67 52.1 0.0 107.4

Manual Integrations




APPROVED


Reviewed By :John Carlone 05/05/2022  
Supervised By :Mahesh Dadoda 05/05/2022



#34  
1,4-Difluorobenzene  
Concen: 50.000 ug/l  
RT: 6.763 min Scan# 931  
Delta R.T. -0.000 min  
Lab File: VX028526.D  
Acq: 04 May 2022 17:26

Tgt Ion:114 Resp: 534796  
Ion Ratio Lower Upper  
114 100  
63 16.6 0.0 36.8  
88 14.8 0.0 32.4





#35

Dibromofluoromethane

Concen: 51.720 ug/l

RT: 5.385 min Scan# 705

Delta R.T. 0.000 min

Lab File: VX028526.D

Acq: 04 May 2022 17:26

Instrument :

MSVOA\_X

ClientSampleId :

VSTDCCC050EC

Tgt Ion: 113 Resp: 193970





| Ion | Ratio | Lower | Upper |
|-----|-------|-------|-------|
| 113 | 100   |       |       |
| 111 | 102.4 | 82.8  | 124.2 |
| 192 | 21.1  | 16.9  | 25.3  |

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/05/2022

Supervised By :Mahesh Dadoda 05/05/2022



#36

1,1-Dichloropropene

Concen: 47.701 ug/l

RT: 5.690 min Scan# 755




Delta R.T. -0.006 min


Lab File: VX028526.D

Acq: 04 May 2022 17:26

Tgt Ion: 75 Resp: 233359


| Ion | Ratio | Lower | Upper |
|-----|-------|-------|-------|
| 75  | 100   |       |       |
| 110 | 38.8  | 18.3  | 54.8  |
| 77  | 31.2  | 24.8  | 37.2  |





#37  
Ethyl Acetate  
Concen: 49.529 ug/l  
RT: 4.708 min Scan# 594  
Delta R.T. 0.000 min  
Lab File: VX028526.D  
Acq: 04 May 2022 17:26




Instrument :  
MSVOA\_X  
ClientSampleId :  
VSTDCCC050EC




Tgt Ion: 43 Resp: 245864  
Ion Ratio Lower Upper  
43 100  
61 15.6 12.6 19.0  
70 12.9 10.2 15.2

Manual Integrations  
APPROVED



Reviewed By :John Carlone 05/05/2022  
Supervised By :Mahesh Dadoda 05/05/2022




#38  
Carbon Tetrachloride  
Concen: 48.627 ug/l  
RT: 5.678 min Scan# 753  
Delta R.T. -0.006 min  
Lab File: VX028526.D  
Acq: 04 May 2022 17:26



Tgt Ion:117 Resp: 268552  
Ion Ratio Lower Upper  
117 100  
119 97.0 77.3 115.9  
121 31.4 24.2 36.2









#39  
Methylcyclohexane  
Concen: 43.840 ug/l  
RT: 7.379 min Scan# 1033  
Delta R.T. -0.006 min  
Lab File: VX028526.D  
Acq: 04 May 2022 17:26

Instrument :  
MSVOA\_X  
ClientSampleId :  
VSTDCCC050EC

Tgt Ion: 83 Resp: 270611  
Ion Ratio Lower Upper  
83 100  
55 64.6 54.6 81.8  
98 50.5 37.4 56.2




Manual Integrations  
APPROVED

Reviewed By :John Carlone 05/05/2022  
Supervised By :Mahesh Dadoda 05/05/2022




#40  
Benzene  
Concen: 49.416 ug/l  
RT: 6.037 min Scan# 812  
Delta R.T. -0.006 min  
Lab File: VX028526.D  
Acq: 04 May 2022 17:26

Tgt Ion: 78 Resp: 707652  
Ion Ratio Lower Upper  
78 100  
77 24.2 19.3 28.9








#41  
Methacrylonitrile  
Concen: 51.609 ug/l  
RT: 4.916 min Scan# 61  
Delta R.T. -0.006 min  
Lab File: VX028526.D  
Acq: 04 May 2022 17:26

Instrument :

MSVOA\_X

ClientSampleId :

VSTDCCC050EC






Tgt Ion: 41 Resp: 132170  
Ion Ratio Lower Upper  
41 100  
39 54.7 43.0 64.4  
67 86.7 70.2 105.4  
52 33.3 26.0 39.0

Manual Integrations


APPROVED

Reviewed By :John Carlone 05/05/2022



Supervised By :Mahesh Dadoda 05/05/2022




#42  
1,2-Dichloroethane  
Concen: 47.781 ug/l  
RT: 6.086 min Scan# 820  
Delta R.T. -0.006 min  
Lab File: VX028526.D  
Acq: 04 May 2022 17:26




Tgt Ion: 62 Resp: 253993  
Ion Ratio Lower Upper  
62 100  
98 10.4 0.0 20.4





#43  
Isopropyl Acetate  
Concen: 49.108 ug/l  
RT: 6.336 min Scan# 80  
Delta R.T. 0.000 min  
Lab File: VX028526.D  
Acq: 04 May 2022 17:26




Instrument :  
MSVOA\_X  
ClientSampleId :  
VSTDCCC050EC



Tgt Ion: 43 Resp: 39674  
Ion Ratio Lower Upper  
43 100  
61 22.2 18.5 27.7  
87 16.7 13.0 19.6




Manual Integrations  
APPROVED


Reviewed By :John Carlone 05/05/2022  
Supervised By :Mahesh Dadoda 05/05/2022



#44  
Trichloroethene  
Concen: 50.937 ug/l  
RT: 7.123 min Scan# 990  
Delta R.T. -0.006 min  
Lab File: VX028526.D  
Acq: 04 May 2022 17:26

Tgt Ion:130 Resp: 211529  
Ion Ratio Lower Upper  
130 100  
95 94.5 0.0 188.8





#45

1,2-Dichloropropane

Concen: 50.808 ug/l

RT: 7.428 min Scan# 1040

Delta R.T. -0.006 min

Lab File: VX028526.D

Acq: 04 May 2022 17:26

Instrument :

MSVOA\_X

ClientSampleId :

VSTDCCC050EC

Tgt Ion: 63 Resp: 17809

Ion Ratio Lower Upper

63 100





65 30.8 25.8 38.6

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/05/2022

Supervised By :Mahesh Dadoda 05/05/2022



#46

Dibromomethane

Concen: 50.431 ug/l

RT: 7.580 min Scan# 1065

Delta R.T. -0.000 min

Lab File: VX028526.D

Acq: 04 May 2022 17:26




Tgt Ion: 93 Resp: 138668


Ion Ratio Lower Upper

93 100

95 82.7 65.8 98.6

174 107.7 85.4 128.2






#47  
Bromodichloromethane  
Concen: 51.116 ug/l  
RT: 7.824 min Scan# 1105  
Delta R.T. -0.000 min  
Lab File: VX028526.D  
Acq: 04 May 2022 17:26

Instrument :

MSVOA\_X

ClientSampleId :

VSTDCCC050EC






Tgt Ion: 83 Resp: 270139  
Ion Ratio Lower Upper  
83 100  
85 63.7 52.2 78.2  
127 9.6 6.8 10.2

Manual Integrations


APPROVED

Reviewed By :John Carlone 05/05/2022



Supervised By :Mahesh Dadoda 05/05/2022




#48  
Methyl methacrylate  
Concen: 50.040 ug/l  
RT: 7.690 min Scan# 1083  
Delta R.T. -0.006 min  
Lab File: VX028526.D  
Acq: 04 May 2022 17:26



Tgt Ion: 41 Resp: 185606  
Ion Ratio Lower Upper  
41 100  
69 98.7 82.5 123.7  
39 63.5 43.8 65.8





#49

1,4-Dioxane

Concen: 877.828 ug/l

RT: 7.659 min Scan# 1078

Delta R.T. 0.000 min

Lab File: VX028526.D

Acq: 04 May 2022 17:26

Instrument :

MSVOA\_X

ClientSampleId :

VSTDCCC050EC

Tgt Ion: 88 Resp: 87104





| Ion | Ratio | Lower | Upper |
|-----|-------|-------|-------|
| 88  | 100   |       |       |
| 43  | 27.1  | 21.9  | 32.9  |
| 58  | 60.8  | 53.7  | 80.5  |

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/05/2022

Supervised By :Mahesh Dadoda 05/05/2022



#50

Toluene-d8

Concen: 47.867 ug/l

RT: 8.647 min Scan# 1240




Delta R.T. -0.006 min


Lab File: VX028526.D

Acq: 04 May 2022 17:26

Tgt Ion: 98 Resp: 689516

| Ion | Ratio | Lower | Upper |
|-----|-------|-------|-------|
| 98  | 100   |       |       |
| 100 | 66.0  | 52.6  | 78.8  |





#51  
4-Methyl-2-Pentanone  
Concen: 250.891 ug/l  
RT: 8.574 min Scan# 1228  
Delta R.T. -0.000 min  
Lab File: VX028526.D  
Acq: 04 May 2022 17:26

Instrument :

MSVOA\_X

ClientSampleId :





VSTDCCC050EC

Tgt Ion: 43 Resp: 1300709  
Ion Ratio Lower Upper  
43 100  
58 43.6 36.1 54.1

Manual Integrations




APPROVED


Reviewed By :John Carlone 05/05/2022  
Supervised By :Mahesh Dadoda 05/05/2022



#52  
Toluene  
Concen: 48.599 ug/l  
RT: 8.720 min Scan# 1252  
Delta R.T. -0.000 min  
Lab File: VX028526.D  
Acq: 04 May 2022 17:26

Tgt Ion: 92 Resp: 483506  
Ion Ratio Lower Upper  
92 100  
91 170.1 137.5 206.3





#53

t-1,3-Dichloropropene

Concen: 49.747 ug/l

RT: 8.976 min Scan# 1194

Delta R.T. -0.006 min

Lab File: VX028526.D

Acq: 04 May 2022 17:26

Instrument :

MSVOA\_X

ClientSampleId :

VSTDCCC050EC

Tgt Ion: 75 Resp: 279409

Ion Ratio Lower Upper

75 100





77 31.1 25.4 38.0

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/05/2022

Supervised By :Mahesh Dadoda 05/05/2022



#54

cis-1,3-Dichloropropene

Concen: 50.417 ug/l

RT: 8.366 min Scan# 1194

Delta R.T. 0.000 min

Lab File: VX028526.D

Acq: 04 May 2022 17:26




Tgt Ion: 75 Resp: 304083


Ion Ratio Lower Upper

75 100

77 31.8 24.8 37.2

39 41.5 29.6 44.4





#55

1,1,2-Trichloroethane

Concen: 52.549 ug/l

RT: 9.153 min Scan# 1

Delta R.T. -0.000 min

Lab File: VX028526.D

Acq: 04 May 2022 17:26

Instrument :

MSVOA\_X

ClientSampleId :

VSTDCCC050EC

Tgt Ion: 97 Resp: 212420

Ion Ratio Lower Upper

97 100

83 81.3 66.2 99.4

85 54.1 42.9 64.3





99 64.4 49.0 73.6

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/05/2022

Supervised By :Mahesh Dadoda 05/05/2022



#56

Ethyl methacrylate

Concen: 53.297 ug/l

RT: 9.116 min Scan# 1317

Delta R.T. 0.000 min

Lab File: VX028526.D

Acq: 04 May 2022 17:26




Tgt Ion: 69 Resp: 312297

Ion Ratio Lower Upper


69 100

41 56.1 43.8 65.6

39 35.5 22.2 33.2#








#57  
1,3-Dichloropropane  
Concen: 50.986 ug/l  
RT: 9.311 min Scan# 11349  
Delta R.T. 0.000 min  
Lab File: VX028526.D  
Acq: 04 May 2022 17:26




Instrument :  
MSVOA\_X  
ClientSampleId :  
VSTDCCC050EC




Tgt Ion: 76 Resp: 332652  
Ion Ratio Lower Upper  
76 100  
78 32.9 25.8 38.8

Manual Integrations  
APPROVED



Reviewed By :John Carlone 05/05/2022  
Supervised By :Mahesh Dadoda 05/05/2022




#58  
2-Chloroethyl Vinyl ether  
Concen: 254.987 ug/l  
RT: 8.244 min Scan# 1174  
Delta R.T. 0.000 min  
Lab File: VX028526.D  
Acq: 04 May 2022 17:26



Tgt Ion: 63 Resp: 701030  
Ion Ratio Lower Upper  
63 100  
106 33.3 23.9 35.9









#59  
2-Hexanone  
Concen: 248.918 ug/l  
RT: 9.433 min Scan# 11  
Delta R.T. 0.000 min  
Lab File: VX028526.D  
Acq: 04 May 2022 17:26

Instrument :  
MSVOA\_X  
ClientSampleId :  
VSTDCCC050EC

Tgt Ion: 43 Resp: 999620  
Ion Ratio Lower Upper  
43 100  
58 60.6 31.9 95.7




Manual Integrations  
**APPROVED**


Reviewed By :John Carlone 05/05/2022  
Supervised By :Mahesh Dadoda 05/05/2022



#60  
Dibromochloromethane  
Concen: 52.539 ug/l  
RT: 9.525 min Scan# 1384  
Delta R.T. -0.000 min  
Lab File: VX028526.D  
Acq: 04 May 2022 17:26

Tgt Ion:129 Resp: 231528  
Ion Ratio Lower Upper  
129 100  
127 78.8 40.3 120.8





#61  
1,2-Dibromoethane  
Concen: 50.749 ug/l  
RT: 9.610 min Scan# 1398  
Delta R.T. 0.000 min  
Lab File: VX028526.D  
Acq: 04 May 2022 17:26

Instrument :

MSVOA\_X

ClientSampleId :





VSTDCCC050EC

Tgt Ion:107 Resp: 22505  
Ion Ratio Lower Upper  
107 100  
109 93.8 74.4 111.6

Manual Integrations




APPROVED


Reviewed By :John Carlone 05/05/2022  
Supervised By :Mahesh Dadoda 05/05/2022



#62  
4-Bromofluorobenzene  
Concen: 49.032 ug/l  
RT: 11.079 min Scan# 1639  
Delta R.T. 0.000 min  
Lab File: VX028526.D  
Acq: 04 May 2022 17:26


Tgt Ion: 95 Resp: 270059  
Ion Ratio Lower Upper  
95 100  
174 88.9 0.0 163.0  
176 85.7 0.0 155.0





#63  
Chlorobenzene-d5  
Concen: 50.000 ug/l  
RT: 10.055 min Scan# 1471  
Delta R.T. 0.000 min  
Lab File: VX028526.D  
Acq: 04 May 2022 17:26




Instrument :  
MSVOA\_X  
ClientSampleId :  
VSTDCCC050EC




Tgt Ion:117 Resp: 505000  
Ion Ratio Lower Upper  
117 100  
82 51.9 41.8 62.8  
119 32.0 24.8 37.2

Manual Integrations  
APPROVED



Reviewed By :John Carlone 05/05/2022  
Supervised By :Mahesh Dadoda 05/05/2022




#64  
Tetrachloroethene  
Concen: 51.717 ug/l  
RT: 9.275 min Scan# 1343  
Delta R.T. -0.000 min  
Lab File: VX028526.D  
Acq: 04 May 2022 17:26




Tgt Ion:164 Resp: 204001  
Ion Ratio Lower Upper  
164 100  
166 126.8 101.2 151.8  
129 93.5 73.0 109.4  
131 87.7 74.2 111.2





#65  
Chlorobenzene  
Concen: 51.771 ug/l  
RT: 10.079 min Scan# 1475  
Delta R.T. 0.000 min  
Lab File: VX028526.D  
Acq: 04 May 2022 17:26




Instrument :  
MSVOA\_X  
ClientSampleId :  
VSTDCCC050EC




Tgt Ion:112 Resp: 55743  
Ion Ratio Lower Upper  
112 100  
114 32.5 25.8 38.8

Manual Integrations  
APPROVED



Reviewed By :John Carlone 05/05/2022  
Supervised By :Mahesh Dadoda 05/05/2022




#66  
1,1,1,2-Tetrachloroethane  
Concen: 54.130 ug/l  
RT: 10.165 min Scan# 1489  
Delta R.T. -0.000 min  
Lab File: VX028526.D  
Acq: 04 May 2022 17:26




Tgt Ion:131 Resp: 217112  
Ion Ratio Lower Upper  
131 100  
133 95.4 48.3 144.8  
119 64.7 32.6 98.0





#67  
Ethyl Benzene  
Concen: 52.469 ug/l  
RT: 10.195 min Scan# 1494  
Delta R.T. 0.000 min  
Lab File: VX028526.D  
Acq: 04 May 2022 17:26




Instrument :  
MSVOA\_X  
ClientSampleId :  
VSTDCCC050EC




Tgt Ion: 91 Resp: 947300  
Ion Ratio Lower Upper  
91 100  
106 32.6 24.6 36.8

Manual Integrations  
APPROVED



Reviewed By :John Carlone 05/05/2022  
Supervised By :Mahesh Dadoda 05/05/2022




#68  
m/p-Xylenes  
Concen: 103.845 ug/l  
RT: 10.305 min Scan# 1512  
Delta R.T. 0.000 min  
Lab File: VX028526.D  
Acq: 04 May 2022 17:26




Tgt Ion:106 Resp: 763061  
Ion Ratio Lower Upper  
106 100  
91 193.1 160.4 240.6





#69  
o-Xylene  
Concen: 51.558 ug/l  
RT: 10.640 min Scan# 1568  
Delta R.T. -0.006 min  
Lab File: VX028526.D  
Acq: 04 May 2022 17:26




Instrument :  
MSVOA\_X  
ClientSampleId :  
VSTDCCC050EC



Tgt Ion:106 Resp: 37772  
Ion Ratio Lower Upper  
106 100  
91 207.5 106.1 318.1




Manual Integrations  
APPROVED


Reviewed By :John Carlone 05/05/2022  
Supervised By :Mahesh Dadoda 05/05/2022



#70  
Styrene  
Concen: 54.058 ug/l  
RT: 10.659 min Scan# 1570  
Delta R.T. -0.000 min  
Lab File: VX028526.D  
Acq: 04 May 2022 17:26


Tgt Ion:104 Resp: 641612  
Ion Ratio Lower Upper  
104 100  
78 47.1 38.6 57.8  
103 54.8 43.8 65.8





#71  
Bromoform  
Concen: 49.339 ug/l  
RT: 10.799 min Scan# 11  
Delta R.T. -0.000 min  
Lab File: VX028526.D  
Acq: 04 May 2022 17:26




Instrument :  
MSVOA\_X  
ClientSampleId :  
VSTDCCC050EC




Tgt Ion:173 Resp: 182424  
Ion Ratio Lower Upper  
173 100  
175 47.5 24.1 72.2  
254 0.1 0.1 0.1

Manual Integrations  
APPROVED



Reviewed By :John Carlone 05/05/2022  
Supervised By :Mahesh Dadoda 05/05/2022




#72  
1,4-Dichlorobenzene-d4  
Concen: 50.000 ug/l  
RT: 12.024 min Scan# 1794  
Delta R.T. 0.000 min  
Lab File: VX028526.D  
Acq: 04 May 2022 17:26



Tgt Ion:152 Resp: 274575  
Ion Ratio Lower Upper  
152 100  
115 80.7 37.5 112.7  
150 175.3 0.0 338.2








#73  
Isopropylbenzene  
Concen: 52.709 ug/l  
RT: 10.963 min Scan# 1620  
Delta R.T. 0.000 min  
Lab File: VX028526.D  
Acq: 04 May 2022 17:26




Instrument :  
MSVOA\_X  
ClientSampleId :  
VSTDCCC050EC




Tgt Ion:105 Resp: 994743  
Ion Ratio Lower Upper  
105 100  
120 27.4 13.5 40.5

Manual Integrations  
**APPROVED**



Reviewed By :John Carlone 05/05/2022  
Supervised By :Mahesh Dadoda 05/05/2022




#74  
N-ethyl acetate  
Concen: 54.811 ug/l  
RT: 10.842 min Scan# 1600  
Delta R.T. -0.006 min  
Lab File: VX028526.D  
Acq: 04 May 2022 17:26




Tgt Ion: 43 Resp: 357533  
Ion Ratio Lower Upper  
43 100  
70 51.1 42.5 63.7  
55 25.9 23.1 34.7  
61 27.0 22.8 34.2





#75  
1,1,2,2-Tetrachloroethane  
Concen: 54.752 ug/l  
RT: 11.213 min Scan# 1661  
Delta R.T. 0.000 min  
Lab File: VX028526.D  
Acq: 04 May 2022 17:26




Instrument :  
MSVOA\_X  
ClientSampleId :  
VSTDCCC050EC




Tgt Ion: 83 Resp: 335315  
Ion Ratio Lower Upper  
83 100  
131 10.9 5.1 15.3  
85 65.6 32.6 97.7

Manual Integrations  
APPROVED



Reviewed By :John Carlone 05/05/2022  
Supervised By :Mahesh Dadoda 05/05/2022




#76  
1,2,3-Trichloropropane  
Concen: 54.013 ug/l m  
RT: 11.238 min Scan# 1665  
Delta R.T. -0.006 min  
Lab File: VX028526.D  
Acq: 04 May 2022 17:26



Tgt Ion: 75 Resp: 291875  
Ion Ratio Lower Upper  
75 100  
77 38.0 19.4 58.2






#77  
Bromobenzene  
Concen: 53.096 ug/l  
RT: 11.201 min Scan# 1659  
Delta R.T. 0.000 min  
Lab File: VX028526.D  
Acq: 04 May 2022 17:26

Instrument :

MSVOA\_X

ClientSampleId :

VSTDCCC050EC






Tgt Ion:156 Resp: 26097  
Ion Ratio Lower Upper  
156 100  
77 139.7 78.5 235.3  
158 96.5 50.2 150.7

Manual Integrations


APPROVED

Reviewed By :John Carlone 05/05/2022



Supervised By :Mahesh Dadoda 05/05/2022




#78  
n-propylbenzene  
Concen: 53.560 ug/l  
RT: 11.305 min Scan# 1676  
Delta R.T. -0.000 min  
Lab File: VX028526.D  
Acq: 04 May 2022 17:26



Tgt Ion: 91 Resp: 1120860  
Ion Ratio Lower Upper  
91 100  
120 24.9 11.6 34.8









#79  
2-Chlorotoluene  
Concen: 51.804 ug/l  
RT: 11.366 min Scan# 1686  
Delta R.T. -0.000 min  
Lab File: VX028526.D  
Acq: 04 May 2022 17:26

Instrument :  
MSVOA\_X  
ClientSampleId :  
VSTDCCC050EC

Tgt Ion: 91 Resp: 674000  
Ion Ratio Lower Upper  
91 100  
126 36.9 17.3 51.9




Manual Integrations  
APPROVED


Reviewed By :John Carlone 05/05/2022  
Supervised By :Mahesh Dadoda 05/05/2022



#80  
1,3,5-Trimethylbenzene  
Concen: 52.066 ug/l  
RT: 11.451 min Scan# 1700  
Delta R.T. 0.000 min  
Lab File: VX028526.D  
Acq: 04 May 2022 17:26


Tgt Ion:105 Resp: 834093  
Ion Ratio Lower Upper  
105 100  
120 50.9 24.3 72.9





#81  
trans-1,4-Dichloro-2-butene  
Concen: 51.191 ug/l  
RT: 11.018 min Scan# 1629  
Delta R.T. 0.000 min  
Lab File: VX028526.D  
Acq: 04 May 2022 17:26




Instrument :  
MSVOA\_X  
ClientSampleId :  
VSTDCCC050EC




Tgt Ion: 75 Resp: 93080  
Ion Ratio Lower Upper  
75 100  
53 96.5 69.0 103.4  
89 49.1 37.0 55.6

Manual Integrations  
APPROVED



Reviewed By :John Carlone 05/05/2022  
Supervised By :Mahesh Dadoda 05/05/2022




#82  
4-Chlorotoluene  
Concen: 51.926 ug/l  
RT: 11.457 min Scan# 1701  
Delta R.T. 0.000 min  
Lab File: VX028526.D  
Acq: 04 May 2022 17:26




Tgt Ion: 91 Resp: 778214  
Ion Ratio Lower Upper  
91 100  
126 32.1 14.9 44.5





#83  
tert-Butylbenzene  
Concen: 52.708 ug/l  
RT: 11.719 min Scan# 1174  
Delta R.T. 0.000 min  
Lab File: VX028526.D  
Acq: 04 May 2022 17:26




Instrument :  
MSVOA\_X  
ClientSampleId :  
VSTDCCC050EC




Tgt Ion:119 Resp: 862232  
Ion Ratio Lower Upper  
119 100  
91 54.3 26.9 80.6  
134 23.7 11.4 34.2

Manual Integrations  
APPROVED



Reviewed By :John Carlone 05/05/2022  
Supervised By :Mahesh Dadoda 05/05/2022




#84  
1,2,4-Trimethylbenzene  
Concen: 52.728 ug/l  
RT: 11.750 min Scan# 1749  
Delta R.T. -0.006 min  
Lab File: VX028526.D  
Acq: 04 May 2022 17:26



Tgt Ion:105 Resp: 839677  
Ion Ratio Lower Upper  
105 100  
120 46.7 22.6 67.8






#85  
 sec-Butylbenzene  
 Concen: 52.539 ug/l  
 RT: 11.890 min Scan# 1772  
 Delta R.T. 0.000 min  
 Lab File: VX028526.D  
 Acq: 04 May 2022 17:26

Instrument :

MSVOA\_X

ClientSampleId :

VSTDCCC050EC





Tgt Ion:105 Resp: 1036898  
 Ion Ratio Lower Upper  
 105 100  
 134 21.2 10.2 30.4

Manual Integrations


APPROVED

Reviewed By :John Carlone 05/05/2022



Supervised By :Mahesh Dadoda 05/05/2022




#86  
 p-Isopropyltoluene  
 Concen: 52.106 ug/l  
 RT: 12.012 min Scan# 1792  
 Delta R.T. 0.000 min  
 Lab File: VX028526.D  
 Acq: 04 May 2022 17:26



Tgt Ion:119 Resp: 887161  
 Ion Ratio Lower Upper  
 119 100  
 134 27.4 13.0 39.0  
 91 22.5 11.3 33.9





#87

1,3-Dichlorobenzene

Concen: 53.405 ug/l

RT: 11.969 min Scan# 1785

Delta R.T. 0.000 min

Lab File: VX028526.D

Acq: 04 May 2022 17:26

Instrument :

MSVOA\_X

ClientSampleId :

VSTDCCC050EC

Tgt Ion:146 Resp: 49322

Ion Ratio Lower Upper

146 100

111 38.3 19.7 59.1





148 63.8 31.6 94.7

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/05/2022

Supervised By :Mahesh Dadoda 05/05/2022



#88

1,4-Dichlorobenzene

Concen: 51.445 ug/l

RT: 12.043 min Scan# 1797

Delta R.T. 0.001 min

Lab File: VX028526.D

Acq: 04 May 2022 17:26




Tgt Ion:146 Resp: 498863

Ion Ratio Lower Upper


146 100

111 37.8 19.2 57.6

148 64.1 32.3 96.9








#89  
n-Butylbenzene  
Concen: 52.713 ug/l  
RT: 12.335 min Scan# 1845  
Delta R.T. 0.000 min  
Lab File: VX028526.D  
Acq: 04 May 2022 17:26




Instrument :  
MSVOA\_X  
ClientSampleId :  
VSTDCCC050EC




Tgt Ion: 91 Resp: 736952  
Ion Ratio Lower Upper  
91 100  
92 53.8 27.8 83.3  
134 28.6 13.3 39.8

Manual Integrations  
APPROVED



Reviewed By :John Carlone 05/05/2022  
Supervised By :Mahesh Dadoda 05/05/2022




#90  
Hexachloroethane  
Concen: 49.389 ug/l  
RT: 12.542 min Scan# 1879  
Delta R.T. 0.000 min  
Lab File: VX028526.D  
Acq: 04 May 2022 17:26




Tgt Ion:117 Resp: 151236  
Ion Ratio Lower Upper  
117 100  
201 80.7 39.5 118.5





#91  
1,2-Dichlorobenzene  
Concen: 53.509 ug/l  
RT: 12.335 min Scan# 1845  
Delta R.T. 0.000 min  
Lab File: VX028526.D  
Acq: 04 May 2022 17:26




Instrument :  
MSVOA\_X  
ClientSampleId :  
VSTDCCC050EC




Tgt Ion:146 Resp: 490917  
Ion Ratio Lower Upper  
146 100  
111 39.9 21.0 63.0  
148 64.5 32.6 97.7

Manual Integrations  
APPROVED



Reviewed By :John Carlone 05/05/2022  
Supervised By :Mahesh Dadoda 05/05/2022




#92  
1,2-Dibromo-3-Chloropropane  
Concen: 51.882 ug/l  
RT: 12.945 min Scan# 1945  
Delta R.T. -0.000 min  
Lab File: VX028526.D  
Acq: 04 May 2022 17:26




Tgt Ion: 75 Resp: 70663  
Ion Ratio Lower Upper  
75 100  
155 102.4 47.0 141.2  
157 129.4 57.9 173.6





#93  
1,2,4-Trichlorobenzene  
Concen: 56.880 ug/l  
RT: 13.591 min Scan# 2051  
Delta R.T. 0.000 min  
Lab File: VX028526.D  
Acq: 04 May 2022 17:26




Instrument :  
MSVOA\_X  
ClientSampleId :  
VSTDCCC050EC




Tgt Ion:180 Resp: 310378  
Ion Ratio Lower Upper  
180 100  
182 95.2 47.7 143.1  
145 30.2 15.2 45.5

Manual Integrations  
APPROVED



Reviewed By :John Carlone 05/05/2022  
Supervised By :Mahesh Dadoda 05/05/2022




#94  
Hexachlorobutadiene  
Concen: 53.057 ug/l  
RT: 13.725 min Scan# 2073  
Delta R.T. 0.000 min  
Lab File: VX028526.D  
Acq: 04 May 2022 17:26




Tgt Ion:225 Resp: 129191  
Ion Ratio Lower Upper  
225 100  
223 63.5 34.2 102.5  
227 64.0 33.9 101.7





#95  
Naphthalene  
Concen: 56.919 ug/l  
RT: 13.774 min Scan# 2081  
Delta R.T. -0.006 min  
Lab File: VX028526.D  
Acq: 04 May 2022 17:26




Instrument :  
MSVOA\_X  
ClientSampleId :  
VSTDCCC050EC




Tgt Ion:128 Resp: 103465  
Ion Ratio Lower Upper  
128 100  
127 13.0 10.2 15.4  
129 10.9 8.7 13.1

Manual Integrations  
APPROVED

Reviewed By :John Carlone 05/05/2022  
Supervised By :Mahesh Dadoda 05/05/2022



#96  
1,2,3-Trichlorobenzene  
Concen: 57.201 ug/l  
RT: 13.963 min Scan# 2112  
Delta R.T. -0.000 min  
Lab File: VX028526.D  
Acq: 04 May 2022 17:26



Tgt Ion:180 Resp: 306378  
Ion Ratio Lower Upper  
180 100  
182 96.4 50.0 149.8  
145 32.4 16.2 48.5

