



Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX050422\
 Data File : VX028517.D
 Acq On : 04 May 2022 09:40
 Operator : JC/MD
 Sample : BFB
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X041922W.M
 Title : SW846 8260
 Last Update : Tue Apr 19 13:38:04 2022



AutoFind: Scans 1638, 1639, 1640; Background Corrected with Scan 1631

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	18.5	14763	PASS
75	95	30	60	50.6	40312	PASS
95	95	100	100	100.0	79715	PASS
96	95	5	9	6.1	4832	PASS
173	174	0.00	2	1.1	684	PASS
174	95	50	100	78.0	62155	PASS
175	174	5	9	7.4	4581	PASS
176	174	95	101	95.6	59408	PASS
177	176	5	9	7.1	4189	PASS