

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX050520\
 Data File : VX016065.D
 Acq On : 05 May 2020 15:43
 Operator : JC/SP
 Sample : L2469-25
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 SA-DUP01-20200428

Quant Time: May 06 09:16:40 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X050420W.M
 Quant Title : SW846 8260
 QLast Update : Wed May 06 06:43:32 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.64	168	295571	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.83	114	480462	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	474281	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	247417	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.03	65	183270	47.12	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.24%	
35) Dibromofluoromethane	5.47	113	141128	46.36	ug/l	0.00
Spiked Amount 50.000			Recovery	=	92.72%	
50) Toluene-d8	8.70	98	595892	50.81	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.62%	
62) 4-Bromofluorobenzene	11.12	95	239299	53.64	ug/l	0.00
Spiked Amount 50.000			Recovery	=	107.28%	
Target Compounds						
6) Chloroethane	1.72	64	5521	2.341	ug/l	97
12) 1,1-Dichloroethene	2.36	96	2755	0.897	ug/l	97
16) Acetone	2.42	43	2846	1.640	ug/l	91
24) 1,1-Dichloroethane	3.67	63	40771	6.858	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX050520\
Data File : VX016065.D
Acq On : 05 May 2020 15:43
Operator : JC/SP
Sample : L2469-25
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
SA-DUP01-20200428

Quant Time: May 06 09:16:40 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X050420W.M
Quant Title : SW846 8260
QLast Update : Wed May 06 06:43:32 2020
Response via : Initial Calibration

