

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX050522\
 Data File : VX028531.D
 Acq On : 05 May 2022 13:18
 Operator : JC/MD
 Sample : VX0505WBL02
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0505WBL02

Quant Time: May 06 03:38:00 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X041922W.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 19 13:38:04 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	403859	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	686233	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	658153	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	318077	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	275014	52.226	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	104.460%
35) Dibromofluoromethane	5.391	113	259854	53.996	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	108.000%
50) Toluene-d8	8.653	98	934237	50.543	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	101.080%
62) 4-Bromofluorobenzene	11.079	95	343563	48.612	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	97.220%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX050522\
Data File : VX028531.D
Acq On : 05 May 2022 13:18
Operator : JC/MD
Sample : VX0505WBL02
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 5 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0505WBL02

Quant Time: May 06 03:38:00 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X041922W.M
Quant Title : SW846 8260
QLast Update : Tue Apr 19 13:38:04 2022
Response via : Initial Calibration

