



Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX050522\
 Data File : VX028527.D
 Acq On : 05 May 2022 09:39
 Operator : JC/MD
 Sample : BFB
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BFB

Integration File: RTEINT3.P

Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\624X042822W.M
 Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 Last Update : Fri Apr 29 06:09:36 2022



AutoFind: Scans 1638, 1639, 1640; Background Corrected with Scan 1632

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	17.3	17392	PASS
75	95	30	60	50.8	51000	PASS
95	95	100	100	100.0	100323	PASS
96	95	5	9	6.5	6472	PASS
173	174	0.00	2	1.1	889	PASS
174	95	50	100	79.9	80171	PASS
175	174	5	9	7.5	6012	PASS
176	174	95	101	97.6	78224	PASS
177	176	5	9	6.3	4958	PASS