

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX050523\
 Data File : VX035527.D
 Acq On : 05 May 2023 11:19
 Operator : JC/MD
 Sample : VX0505WBL01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0505WBL01

Quant Time: May 08 00:49:49 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X042523W.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 25 16:31:40 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	282444	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	474490	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	424338	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	193278	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	171896	50.446	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	= 100.900%	
35) Dibromofluoromethane	5.385	113	147976	50.811	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 101.620%	
50) Toluene-d8	8.647	98	551992	50.245	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	= 100.480%	
62) 4-Bromofluorobenzene	11.079	95	204959	49.594	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 99.180%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX050523\
Data File : VX035527.D
Acq On : 05 May 2023 11:19
Operator : JC/MD
Sample : VX0505WBL01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0505WBL01

Quant Time: May 08 00:49:49 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X042523W.M
Quant Title : SW846 8260
QLast Update : Tue Apr 25 16:31:40 2023
Response via : Initial Calibration

