

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX050523\
 Data File : VX035532.D
 Acq On : 05 May 2023 13:42
 Operator : JC/MD
 Sample : 02615-02
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 002-1

Quant Time: May 08 00:50:41 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X042523W.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 25 16:31:40 2023
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	269007	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	448313	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	408226	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	191037	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	165528	51.004	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	102.000%
35) Dibromofluoromethane	5.385	113	140824	51.179	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	102.360%
50) Toluene-d8	8.647	98	526428	50.716	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	101.440%
62) 4-Bromofluorobenzene	11.079	95	198361	50.800	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	101.600%
Target Compounds						
						Qvalue
30) Chloroform	5.093	83	67711	13.087	ug/l	99
47) Bromodichloromethane	7.824	83	29129	7.409	ug/l #	94
52) Toluene	8.726	92	1952	0.250	ug/l	100
60) Dibromochloromethane	9.525	129	10191	3.415	ug/l	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX050523\
Data File : VX035532.D
Acq On : 05 May 2023 13:42
Operator : JC/MD
Sample : 02615-02
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
002-1

Quant Time: May 08 00:50:41 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X042523W.M
Quant Title : SW846 8260
QLast Update : Tue Apr 25 16:31:40 2023
Response via : Initial Calibration

